



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86231

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/08/2017	11:30
2	CAL2	1	Water			7.00	03/08/2017	11:31
3	CAL3	1	Water			10.04	03/08/2017	11:32
4	ICV	1	Water			6.99	03/08/2017	11:33
5	CCV1	1	Water			2.02	03/08/2017	11:34
6	I2172-01	1	Solid	20.02	20	7.14	03/08/2017	11:35
7	I2172-01DUP	1	Solid	20.03	20	7.15	03/08/2017	11:36
8	I2172-03	1	Solid	20.04	20	7.00	03/08/2017	11:37
9	I2172-05	1	Solid	20.06	20	6.51	03/08/2017	11:38
10	I2172-07	1	Solid	20.05	20	6.56	03/08/2017	11:39
11	I2172-09	1	Solid	20.00	20	6.48	03/08/2017	11:40
12	I2172-11	1	Solid	20.04	20	6.54	03/08/2017	11:41
13	I2172-13	1	Solid	20.01	20	6.46	03/08/2017	11:42
14	I2172-15	1	Solid	20.01	20	6.51	03/08/2017	11:43
15	I2172-17	1	Solid	20.03	20	6.42	03/08/2017	11:44
16	CCV2	1	Water			12.02	03/08/2017	11:45
17	I2172-19	1	Solid	20.04	20	6.93	03/08/2017	11:46
18	I2174-01	1	Solid	20.02	20	6.71	03/08/2017	11:47
19	I2174-03	1	Solid	20.02	20	6.90	03/08/2017	11:48
20	I2174-05	1	Solid	20.06	20	6.48	03/08/2017	11:49

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I2174-07	1	Solid	20.05	20	6.33	03/08/2017	11:50
22	I2174-09	1	Solid	20.02	20	6.46	03/08/2017	11:51
23	I2174-11	1	Solid	20.04	20	7.02	03/08/2017	11:52
24	I2174-13	1	Solid	20.02	20	6.50	03/08/2017	11:53
25	I2174-15	1	Solid	20.04	20	6.51	03/08/2017	11:54
26	I2174-17	1	Solid	20.06	20	6.43	03/08/2017	11:55
27	CCV3	1	Water			2.02	03/08/2017	11:56
28	I2174-17DUP	1	Solid	20.05	20	6.44	03/08/2017	11:57
29	CCV4	1	Water			12.01	03/08/2017	11:59

WORKLIST(Hardcopy Internal Chain)

19 86231

WorkList Name : PH I2172

WorkList ID : 96311

Date : 3/8/2017 11:08:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2172-01	pH	Cool 4 deg C	SEVE02	K61	D-38	03/07/2017	9045C
	Solid	I2172-03	pH	Cool 4 deg C	SEVE02	K61	D-41	03/07/2017	9045C
	Solid	I2172-05	pH	Cool 4 deg C	SEVE02	K61	D-44	03/07/2017	9045C
	Solid	I2172-07	pH	Cool 4 deg C	SEVE02	K61	D-47	03/07/2017	9045C
	Solid	I2172-09	pH	Cool 4 deg C	SEVE02	K61	D-49	03/07/2017	9045C
	Solid	I2172-11	pH	Cool 4 deg C	SEVE02	K61	D-51	03/07/2017	9045C
	Solid	I2172-13	pH	Cool 4 deg C	SEVE02	K61	D-53	03/07/2017	9045C
	Solid	I2172-15	pH	Cool 4 deg C	SEVE02	K61	D-55	03/07/2017	9045C
	Solid	I2172-17	pH	Cool 4 deg C	SEVE02	K61	D-57	03/07/2017	9045C
	Solid	I2172-19	pH	Cool 4 deg C	SEVE02	K61	D-61	03/07/2017	9045C
	Solid	I2174-01	pH	Cool 4 deg C	SEVE02	K61	D-64	03/07/2017	9045C
	Solid	I2174-03	pH	Cool 4 deg C	SEVE02	K61	D-67	03/07/2017	9045C
	Solid	I2174-05	pH	Cool 4 deg C	SEVE02	K61	D-69	03/07/2017	9045C
	Solid	I2174-07	pH	Cool 4 deg C	SEVE02	K61	D-71	03/07/2017	9045C
	Solid	I2174-09	pH	Cool 4 deg C	SEVE02	K61	D-74	03/07/2017	9045C
	Solid	I2174-11	pH	Cool 4 deg C	SEVE02	K61	D-77	03/07/2017	9045C
	Solid	I2174-13	pH	Cool 4 deg C	SEVE02	K61	D-79	03/07/2017	9045C
	Solid	I2174-15	pH	Cool 4 deg C	SEVE02	K61	D-81	03/07/2017	9045C
	Solid	I2174-17	pH	Cool 4 deg C	SEVE02	K61	D-85	03/07/2017	9045C

Date/Time 03/08/17 11:08 AM
 Received by: 711
 Relinquished by: [Signature]

Date/Time 03/08/17 6:30 PM
 Received by: [Signature]
 Relinquished by: JP