


ISM02.4

LB 86232
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45531



A= 0.0000e+000

B= 2.1941e-004

C= -2.8160e-002

Rho= 0.9999775

Accept=Accepted

dope
x=1/milecept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	117	0.000	117					-
0.2	0.200	0.189	-0.011	989	0.0 %	989					-6
2.5	2.500	2.492	-0.008	11487	0.0 %	11487					0
5.0	5.000	5.019	0.019	23003	0.0 %	23003					0
7.5	7.500	7.541	0.041	34497	0.0 %	34497					0
10.0	10.000	9.962	-0.038	45531	0.0 %	45531					0

0.03
318117

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	117	-		0 ISM02.4	PPB	3/8/2017 13:50	Std
	0.2 S0.2	989	-		0.2 ISM02.4	PPB	3/8/2017 13:53	Std
	2.5 S2.5	11487	-		2.5 ISM02.4	PPB	3/8/2017 13:55	Std
	5 S5	23003	-		5 ISM02.4	PPB	3/8/2017 13:57	Std
	7.5 S7.5	34497	-		7.5 ISM02.4	PPB	3/8/2017 14:00	Std
	10 S10	45531	-		10 ISM02.4	PPB	3/8/2017 14:02	Std
ICV	ICV	17029	3.7082 -		ISM02.4	PPB	3/8/2017 14:05	SMPL
ICB	ICB	46	-0.0181 -		ISM02.4	PPB	3/8/2017 14:07	SMPL
CCV057	CCV057	22102	4.8212 -		ISM02.4	PPB	3/8/2017 14:09	SMPL
CCB057	CCB057	44	-0.0185 -		ISM02.4	PPB	3/8/2017 14:12	SMPL
PB97411BL	PBS004	77	-0.0113 -		ISM02.4	PPB	3/8/2017 14:14	SMPL
I2060-01	MCOAC2	1473	0.295 -		ISM02.4	PPB	3/8/2017 14:16	SMPL
I2060-02	MCOAC3	3681	0.7795 -		ISM02.4	PPB	3/8/2017 14:18	SMPL
I2060-03	MCOAC4	5055	1.081 -		ISM02.4	PPB	3/8/2017 14:21	SMPL
I2060-04	MCOAC5	4689	1.0006 -		ISM02.4	PPB	3/8/2017 14:23	SMPL
I2060-05	MCOAC6	19174	4.1788 -		ISM02.4	PPB	3/8/2017 14:25	SMPL
I2060-06	MCOAC7	2738	0.5726 -		ISM02.4	PPB	3/8/2017 14:28	SMPL
I2060-07	MCOAC8	3049	0.6408 -		ISM02.4	PPB	3/8/2017 14:30	SMPL
I2060-08	MCOAC9	1292	0.2553 -		ISM02.4	PPB	3/8/2017 14:32	SMPL
I2060-09	MCOAC9D	1193	0.2336 -		ISM02.4	PPB	3/8/2017 14:34	SMPL
I2060-10	MCOAC9S	12314	2.6736 -		ISM02.4	PPB	3/8/2017 14:37	SMPL
I2060-11	MCOAD0	17212	3.7483 -		ISM02.4	PPB	3/8/2017 14:39	SMPL
I2060-12	MCOAD4	6448	1.3866 -		ISM02.4	PPB	3/8/2017 14:41	SMPL
I2060-13	MCOAD5	1132	0.2202 -		ISM02.4	PPB	3/8/2017 14:43	SMPL
I2060-14	MCOAD6	801	0.1476 -		ISM02.4	PPB	3/8/2017 14:46	SMPL
I2060-15	MCOAD7	1907	0.3903 -		ISM02.4	PPB	3/8/2017 14:48	SMPL
I2060-16	MCOAF7	8451	1.8261 -		ISM02.4	PPB	3/8/2017 14:50	SMPL
I2060-17	MCOAK3	6562	1.4116 -		ISM02.4	PPB	3/8/2017 14:53	SMPL
I2060-18	MCOAK4	7274	1.5678 -		ISM02.4	PPB	3/8/2017 14:55	SMPL
I2060-19	MCOAK5	27820	6.0758 -		ISM02.4	PPB	3/8/2017 14:57	SMPL
CCV058	CCV058	22613	4.9333 -		ISM02.4	PPB	3/8/2017 14:59	SMPL
CCB058	CCB058	66	-0.0137 -		ISM02.4	PPB	3/8/2017 15:02	SMPL
I2060-20	MCOAK6	114192	25.0266 -		ISM02.4	PPB	3/8/2017 15:04	SMPL
I2060-21	MCOAK7	34101	7.4539 -		ISM02.4	PPB	3/8/2017 15:06	SMPL
I2060-22	MCOAK8	17479	3.8069 -		ISM02.4	PPB	3/8/2017 15:09	SMPL
PB97412BL	PBS004	53	-0.0165 -		ISM02.4	PPB	3/8/2017 15:11	SMPL
I2068-01	MCOAC2	13191	2.8661 -		ISM02.4	PPB	3/8/2017 15:13	SMPL
I2068-02	MCOAG4	1566	0.3154 -		ISM02.4	PPB	3/8/2017 15:15	SMPL
I2068-03	MCOAG5	7007	1.5092 -		ISM02.4	PPB	3/8/2017 15:18	SMPL
I2068-04	MCOAG5D	6612	1.4226 -		ISM02.4	PPB	3/8/2017 15:20	SMPL
I2068-05	MCOAG5S	17700	3.8554 -		ISM02.4	PPB	3/8/2017 15:25	SMPL
I2068-06	MCOAG7	46211	10.111 -		ISM02.4	PPB	3/8/2017 15:27	SMPL
I2068-07	MCOAG9	3799	0.8054 -		ISM02.4	PPB	3/8/2017 15:30	SMPL
I2068-08	MCOAH0	2466	0.5129 -		ISM02.4	PPB	3/8/2017 15:32	SMPL
I2068-09	MCOAH1	3269	0.6891 -		ISM02.4	PPB	3/8/2017 15:34	SMPL
I2068-10	MCOAH3	1558	0.3137 -		ISM02.4	PPB	3/8/2017 15:37	SMPL

INSTRUMENT ID : CV1

I2068-11	MCOAH5	1944	0.3984 -	ISM02.4	PPB	3/8/2017 15:39	SMPL
I2068-12	MCOAH7	2828	0.5923 -	ISM02.4	PPB	3/8/2017 15:41	SMPL
I2068-13	MCOAH9	1049	0.202 -	ISM02.4	PPB	3/8/2017 15:43	SMPL
I2068-14	MCOAJ0	578	0.0987 -	ISM02.4	PPB	3/8/2017 15:46	SMPL
I2068-15	MCOAA0	3523	0.7448 -	ISM02.4	PPB	3/8/2017 15:48	SMPL
I2068-16	MCOAJ1	33205	7.2573 -	ISM02.4	PPB	3/8/2017 15:50	SMPL
CCV059	CCV059	22695	4.9513 -	ISM02.4	PPB	3/8/2017 15:53	SMPL
CCB059	CCB059	60	-0.015 -	ISM02.4	PPB	3/8/2017 15:55	SMPL
I2068-17	MCOAJ2	20105	4.3831 -	ISM02.4	PPB	3/8/2017 15:57	SMPL
I2068-18	MCOAJ3	27349	5.9725 -	ISM02.4	PPB	3/8/2017 15:59	SMPL
I2068-19	MCOAJ4	36996	8.0891 -	ISM02.4	PPB	3/8/2017 16:02	SMPL
I2068-20	MCOAJ5	30255	6.6101 -	ISM02.4	PPB	3/8/2017 16:04	SMPL
I2068-21	MCOAJ6	582	0.0995 -	ISM02.4	PPB	3/8/2017 16:06	SMPL
I2068-22	MCOAJ7	137408	30.1204 -	ISM02.4	PPB	3/8/2017 16:08	SMPL
PB97413BL	PBS004	-98	-0.0497 -	ISM02.4	PPB	3/8/2017 16:11	SMPL
I2105-01	MCOAA5	5076	1.0856 -	ISM02.4	PPB	3/8/2017 16:13	SMPL
I2105-02	MCOAA5D	4175	0.8879 -	ISM02.4	PPB	3/8/2017 16:15	SMPL
I2105-03	MCOAA5S	16485	3.5888 -	ISM02.4	PPB	3/8/2017 16:18	SMPL
I2105-04	MCOAJ3	268662	58.9187 -	ISM02.4	PPB	3/8/2017 16:20	SMPL
I2105-05	MCOAJ4	25390	5.5426 -	ISM02.4	PPB	3/8/2017 16:22	SMPL
I2105-06	MCOAJ5	335652	HIGH -	ISM02.4	PPB	3/8/2017 16:25	
I2105-07	MCOAJ6	481370	HIGH -	ISM02.4	PPB	3/8/2017 16:26	
I2105-08	MCOAJ7	4597	0.9805 -	ISM02.4	PPB	3/8/2017 16:28	SMPL
I2105-09	MCOAJ8	10495	2.2745 -	ISM02.4	PPB	3/8/2017 16:31	SMPL
I2105-10	MCOAJ9	107091	23.4686 -	ISM02.4	PPB	3/8/2017 16:33	SMPL
I2105-11	MCOAK0	94925	20.7992 -	ISM02.4	PPB	3/8/2017 16:35	SMPL
I2105-12	MCOAK1	86285	18.9036 -	ISM02.4	PPB	3/8/2017 16:37	SMPL
I2105-13	MCOAK2	16375	3.5647 -	ISM02.4	PPB	3/8/2017 16:40	SMPL
CCV060	CCV060	22740	4.9612 -	ISM02.4	PPB	3/8/2017 16:42	SMPL
CCB060	CCB060	68	-0.0132 -	ISM02.4	PPB	3/8/2017 16:44	SMPL
I2105-14	MCOAK9	10992	2.3836 -	ISM02.4	PPB	3/8/2017 16:47	SMPL
I2105-15	MCOAL0	4367	0.93 -	ISM02.4	PPB	3/8/2017 16:49	SMPL
I2105-16	MCOAL1	6862	1.4774 -	ISM02.4	PPB	3/8/2017 16:51	SMPL
I2105-17	MCOAL3	28290	6.1789 -	ISM02.4	PPB	3/8/2017 16:53	SMPL
I2060-20DLX5	MCOAK6	22741	4.9614 -	ISM02.4	PPB	3/8/2017 16:56	SMPL
I2068-06DLX2	MCOAG7	23468	5.1209 -	ISM02.4	PPB	3/8/2017 16:58	SMPL
I2068-22DLX5	MCOAJ7	28871	6.3064 -	ISM02.4	PPB	3/8/2017 17:00	SMPL
I2105-04DLX10	MCOAJ3	28578	6.2421 -	ISM02.4	PPB	3/8/2017 17:03	SMPL
I2105-06DLX10	MCOAJ5	44997	9.8446 -	ISM02.4	PPB	3/8/2017 17:05	SMPL
I2105-07DLX10	MCOAJ6	74183	16.2483 -	ISM02.4	PPB	3/8/2017 17:07	SMPL
I2105-10DLX5	MCOAJ9	22087	4.8179 -	ISM02.4	PPB	3/8/2017 17:09	SMPL
I2105-11DLX5	MCOAK0	18779	4.0921 -	ISM02.4	PPB	3/8/2017 17:12	SMPL
I2105-12DLX5	MCOAK1	17538	3.8198 -	ISM02.4	PPB	3/8/2017 17:14	SMPL
I2105-07DL2X50	MCOAJ6	14431	3.1381 -	ISM02.4	PPB	3/8/2017 17:16	SMPL
CCV061	CCV061	22913	4.9992 -	ISM02.4	PPB	3/8/2017 17:19	SMPL
CCB061	CCB061	63	-0.0143 -	ISM02.4	PPB	3/8/2017 17:21	SMPL