


7470A

LB86248
INSTRUMENT ID : CV1

Linear

 μ Abs.:

34823



A= 0.0000e+000

B= 2.8500e-004

C= -4.3512e-002

Rho= 0.9997398

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	54	0.000	54				
0.2	0.200	0.177	-0.023	775	0.0 %	775				
2.5	2.500	2.537	0.037	9054	0.0 %	9054				
5.0	5.000	4.975	-0.025	17608	0.0 %	17608				
7.5	7.500	7.658	0.158	27024	0.0 %	27024				
10.0	10.000	9.881	-0.119	34823	0.0 %	34823				

LB86248 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Mar 2017 08:34:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	09 Mar 2017 09:03:34	S	Std
0.2 - 1	50.2	775	-	0.2000	7470A	PPB	09 Mar 2017 09:05:42	S	Std
2.5 - 1	52.5	9054	-	2.5000	7470A	PPB	09 Mar 2017 09:07:50	S	Std
5.0 - 1	55	17608	-	5.0000	7470A	PPB	09 Mar 2017 09:09:59	S	Std
7.5 - 1	57.5	27024	-	7.5000	7470A	PPB	09 Mar 2017 09:12:07	S	Std
10.0 - 1	510	34823	-	10.0000	7470A	PPB	09 Mar 2017 09:16:45	S	Std
ICV63 - 1	ICV63	12789	3.6013	-	7470A	PPB	09 Mar 2017 09:19:39	U	SMPL
ICB63 - 1	ICB63	25	-0.0364	-	7470A	PPB	09 Mar 2017 09:21:46	U	SMPL
CCV62 - 1	CCV62	16647	4.7008	-	7470A	PPB	09 Mar 2017 09:28:39	U	SMPL
CCB62 - 1	CCB62	31	-0.0347	-	7470A	PPB	09 Mar 2017 09:30:45	U	SMPL
CRA - 1	CRA	661	0.1449	-	7470A	PPB	09 Mar 2017 09:32:51	U	SMPL
HighStd - 1	HighStd	35354	10.0323	-	7470A	PPB	09 Mar 2017 09:38:45	U	SMPL
ChkStd - 1	ChkStd	23139	6.5510	-	7470A	PPB	09 Mar 2017 09:40:52	U	SMPL
PB97447BL - 1	PBW	4	-0.0424	-	7470A	PPB	09 Mar 2017 09:43:00	U	SMPL
PB97447BS - 1	LCSW	11540	3.2454	-	7470A	PPB	09 Mar 2017 09:45:08	U	SMPL
PB97447TB - 1	PB97409TB	65	-0.0250	-	7470A	PPB	09 Mar 2017 09:47:15	U	SMPL
I2127-22 - 1	D-1	106	-0.0133	-	7470A	PPB	09 Mar 2017 09:49:23	U	SMPL
I2127-22DUP - 1	D-1DUP	169	0.0047	-	7470A	PPB	09 Mar 2017 09:51:32	U	SMPL
I2127-22MS - 1	D-1MS	12001	3.3767	-	7470A	PPB	09 Mar 2017 09:53:40	U	SMPL
I2127-22MSD - 1	D-1MSD	11679	3.2850	-	7470A	PPB	09 Mar 2017 09:55:46	U	SMPL
CCV63 - 1	CCV63	17312	4.8904	-	7470A	PPB	09 Mar 2017 10:00:22	U	SMPL
CCB63 - 1	CCB63	62	-0.0258	-	7470A	PPB	09 Mar 2017 10:03:50	U	SMPL
I2127-23 - 1	D-8	319	0.0474	-	7470A	PPB	09 Mar 2017 10:05:56	U	SMPL
I2127-24 - 1	D-12	227	0.0212	-	7470A	PPB	09 Mar 2017 10:08:02	U	SMPL
I2127-25 - 1	D-14	312	0.0454	-	7470A	PPB	09 Mar 2017 10:10:09	U	SMPL
I2127-26 - 1	D-18	247	0.0269	-	7470A	PPB	09 Mar 2017 10:12:16	U	SMPL
I2127-27 - 1	D-21	153	0.0001	-	7470A	PPB	09 Mar 2017 10:14:24	U	SMPL
I2127-28 - 1	D-23	233	0.0229	-	7470A	PPB	09 Mar 2017 10:16:31	U	SMPL
I2127-29 - 1	D-25	265	0.0320	-	7470A	PPB	09 Mar 2017 10:18:40	U	SMPL
I2127-30 - 1	D-26	330	0.0505	-	7470A	PPB	09 Mar 2017 10:20:48	U	SMPL
I2129-09 - 1	D-27	434	0.0802	-	7470A	PPB	09 Mar 2017 10:22:56	U	SMPL
I2129-10 - 1	D-30	285	0.0377	-	7470A	PPB	09 Mar 2017 10:25:03	U	SMPL
CCV64 - 1	CCV64	18235	5.1534	-	7470A	PPB	09 Mar 2017 10:27:10	U	SMPL
CCB64 - 1	CCB64	64	-0.0253	-	7470A	PPB	09 Mar 2017 10:29:17	U	SMPL
I2129-11 - 1	D-33	328	0.0500	-	7470A	PPB	09 Mar 2017 10:31:23	U	SMPL
I2129-12 - 1	D-35	246	0.0266	-	7470A	PPB	09 Mar 2017 10:33:30	U	SMPL
I2127-22LX5 - 1		125	-0.0079	-	7470A	PPB	09 Mar 2017 10:35:37	U	SMPL
I2127-22A - 1		14258	4.0200	-	7470A	PPB	09 Mar 2017 10:37:44	U	SMPL
CCV65 - 1	CCV65	18320	5.1776	-	7470A	PPB	09 Mar 2017 10:39:52	U	SMPL
CCB65 - 1	CCB65	6	-0.0418	-	7470A	PPB	09 Mar 2017 10:41:59	U	SMPL