

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 03/09/2017

Run Number: LB86250

BalanceID: WC SC-3

DATE & TIME IN: 03/09/2017 12:10

DATE & TIME OUT: 03/09/2017 16:20

OVEN TEMP: 105 °C

OVEN TEMP: 104 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86250BL	LB86250BL	1.5831	1.5831	100	1.5831	1.5831	1.5831	0	0
2	LB86250BS	LB86250BS	1.4843	1.4843	100	1.539	1.539	1.539	0.0547	547
3	I2159-01	EFF-WW	1.4205	1.4205	200	1.4333	1.4333	1.4333	0.0128	64
4	I2171-01	DRAINWATER-TANK-1	1.4111	1.4111	30	1.4235	1.4235	1.4235	0.0124	413.3
5	I2177-01	DRGFB-WQ-068-AD-A	1.4206	1.4206	400	1.4346	1.4346	1.4346	0.014	35
6	I2177-02	DRGFB-WQ-069-BG-A	1.3995	1.3995	400	1.4096	1.4096	1.4096	0.0101	25.3
7	I2183-01	001M-WILLETS-PTBVLD (MAR)	1.4047	1.4047	300	1.4237	1.4237	1.4237	0.019	63.3
8	I2183-02	002M-35THAVE (MAR)	1.4098	1.4098	300	1.4269	1.4269	1.4269	0.0171	57
9	I2185-01	EFF-170307-02	1.4096	1.4096	800	1.4108	1.4108	1.4108	0.0012	1.5
10	I2187-02	OUTFALL001-170307	1.4162	1.4162	600	1.4163	1.4163	1.4163	0.0001	0.2
11	I2190-01	030617-DSDC-E-D-M	1.4196	1.4196	400	1.4403	1.4403	1.4403	0.0207	51.8
12	I2190-02	030617-DSDC-E-U-M	1.4026	1.4026	400	1.4226	1.4226	1.4226	0.02	50
13	I2190-03	030617-TURB-E-D-M	1.4168	1.4168	400	1.441	1.441	1.441	0.0242	60.5
14	I2190-04	030617-TURB-E-U-M	1.4132	1.4132	400	1.4392	1.4392	1.4392	0.026	65
15	I2191-01	030717-TURB-E-D-M	1.4043	1.4043	400	1.4359	1.4359	1.4359	0.0316	79
16	I2191-02	030717-TURB-E-U-M	1.408	1.408	200	1.4153	1.4153	1.4153	0.0073	36.5
17	I2191-02DUP	030717-TURB-E-U-MDUP	1.3969	1.3969	200	1.4045	1.4045	1.4045	0.0076	38
18	I2193-01	COMP	1.3971	1.3971	100	1.4183	1.4183	1.4183	0.0212	212

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
A = Sample Volume (ml) B = Empty Dish Weight (g) C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g) D = Weight (g)										

WORKLIST(Hardcopy Internal Chain)

1886250

WorkList Name : TSS I2183

WorkList ID : 96352

Date : 3/9/2017 8:35:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2159-01 -B	TSS -B	Cool 4 deg C	ARDM01	K12	EFF-VWV	03/07/2017	SM2540 D
	Water	I2171-01	TSS	Cool 4 deg C	MAJO01	K13	DRAINWATER-TANK-1	03/07/2017	SM2540 D
	Water	I2177-01	TSS -A	Cool 4 deg C	JAYC01	K32	DRGFB-WQ-068-AD-A	03/07/2017	SM2540 D
	Water	I2177-02	TSS -A	Cool 4 deg C	JAYC01	K32	DRGFB-WQ-069-BG-A	03/07/2017	SM2540 D
	Water	I2183-01	TSS -E	Cool 4 deg C	TULL01	K11	001M-WILLETS-PTBVLD(MAI	03/07/2017	SM2540 D
	Water	I2183-02	TSS -E	Cool 4 deg C	TULL01	K11	002M-35THAVE(MAR)	03/07/2017	SM2540 D
	Water	I2185-01	TSS -E	Cool 4 deg C	VIAS01	J11	EFF-170307-02	03/07/2017	SM2540 D
	Water	I2187-02	TSS -E	Cool 4 deg C	SHAW22	K11	OUTFALL001-170307	03/07/2017	SM2540 D
	Water	I2190-01 -I	TSS -B	Cool 4 deg C	HDRI03	K62	030617-DSDC-E-D-M	03/06/2017	SM2540 D
	Water	I2190-02	TSS -B	Cool 4 deg C	HDRI03	K62	030617-DSDC-E-U-M	03/06/2017	SM2540 D
	Water	I2190-03	TSS	Cool 4 deg C	HDRI03	K62	030617-TURB-E-D-M	03/06/2017	SM2540 D
	Water	I2190-04	TSS	Cool 4 deg C	HDRI03	K62	030617-TURB-E-U-M	03/06/2017	SM2540 D
	Water	I2191-01	TSS	Cool 4 deg C	HDRI03	K62	030717-TURB-E-D-M	03/07/2017	SM2540 D
	Water	I2191-02	TSS	Cool 4 deg C	HDRI03	K62	030717-TURB-E-U-M	03/07/2017	SM2540 D
	Water	I2193-01	TSS	Cool 4 deg C	ARAM01	K13	COMP	03/08/2017	SM2540 D

Date/Time 03/09/17 8:40 AM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 03/09/17 1:20 PM
 Received by: [Signature]
 Relinquished by: [Signature]