


ISM02.4

LB 86258
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47953



A= 0.0000e+000

B= 2.0839e-004

C= -1.9419e-002

Rho= 0.9999832

Accept=Accepted

slope
y=intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	0.010	0.010	143	0.000	143					-
0.2	0.200	0.192	-0.008	1016	0.0 %	1016					-4
2.5	2.500	2.480	-0.020	11995	0.0 %	11995					-1
5.0	5.000	5.005	0.005	24113	0.0 %	24113					0
7.5	7.500	7.538	0.038	36266	0.0 %	36266					1
10.0	10.000	9.974	-0.026	47953	0.0 %	47953					0

NB
3/9/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	143	-		0 ISM02.4	PPB	3/9/2017 10:58	Std
	0.2 S0.2	1016	-		0.2 ISM02.4	PPB	3/9/2017 11:01	Std
	2.5 S2.5	11995	-		2.5 ISM02.4	PPB	3/9/2017 11:03	Std
	5 S5	24113	-		5 ISM02.4	PPB	3/9/2017 11:05	Std
	7.5 S7.5	36266	-		7.5 ISM02.4	PPB	3/9/2017 11:07	Std
	10 S10	47953	-		10 ISM02.4	PPB	3/9/2017 11:10	Std
ICV	ICV	18894	3.9179 -		ISM02.4	PPB	3/9/2017 11:13	SMPL
ICB	ICB	79	-0.003 -		ISM02.4	PPB	3/9/2017 11:15	SMPL
CCV066	CCV066	24246	5.0332 -		ISM02.4	PPB	3/9/2017 11:17	SMPL
CCB066	CCB066	73	-0.0042 -		ISM02.4	PPB	3/9/2017 11:19	SMPL
PB97451BL	PBS004	85	-0.0017 -		ISM02.4	PPB	3/9/2017 11:22	SMPL
I2084-19	MX2883	19929	4.1336 -		ISM02.4	PPB	3/9/2017 11:24	SMPL
I2084-20	MX2883D	18610	3.8587 -		ISM02.4	PPB	3/9/2017 11:26	SMPL
I2084-21	MX2883S	31139	6.4696 -		ISM02.4	PPB	3/9/2017 11:28	SMPL
PB97452BL	PBS004	0	-0.0194 -		ISM02.4	PPB	3/9/2017 11:31	SMPL
I2104-01	MC0AJ8	153685	32.0071 -		ISM02.4	PPB	3/9/2017 11:33	SMPL
I2104-02	MC0AJ9	9003	1.8567 -		ISM02.4	PPB	3/9/2017 11:35	SMPL
I2104-03	MC0AK0	1116	0.2131 -		ISM02.4	PPB	3/9/2017 11:38	SMPL
I2104-04	MC0AK0D	1061	0.2017 -		ISM02.4	PPB	3/9/2017 11:42	SMPL
I2104-05	MC0AK0S	13814	2.8593 -		ISM02.4	PPB	3/9/2017 11:45	SMPL
I2104-06	MC0AK1	18950	3.9296 -		ISM02.4	PPB	3/9/2017 11:47	SMPL
I2104-07	MC0AK2	23445	4.8663 -		ISM02.4	PPB	3/9/2017 11:49	SMPL
I2104-08	MC0AK3	21924	4.5493 -		ISM02.4	PPB	3/9/2017 11:52	SMPL
I2104-09	MC0AK4	1547	0.303 -		ISM02.4	PPB	3/9/2017 11:54	SMPL
I2104-10	MC0AK5	2150	0.4286 -		ISM02.4	PPB	3/9/2017 11:56	SMPL
I2104-11	MC0AK6	868	0.1615 -		ISM02.4	PPB	3/9/2017 11:58	SMPL
PB97453BL	PBS004	96	0.0006 -		ISM02.4	PPB	3/9/2017 12:01	SMPL
I2151-04	MDABX7	471	0.0787 -		ISM02.4	PPB	3/9/2017 12:05	SMPL
I2151-05	MDABX7D	482	0.081 -		ISM02.4	PPB	3/9/2017 12:07	SMPL
CCV067	CCV067	24346	5.0541 -		ISM02.4	PPB	3/9/2017 12:10	SMPL
CCB067	CCB067	81	-0.0025 -		ISM02.4	PPB	3/9/2017 12:12	SMPL
I2151-06	MDABX7S	12890	2.6667 -		ISM02.4	PPB	3/9/2017 12:14	SMPL
I2151-07	MDABX9	167	0.0154 -		ISM02.4	PPB	3/9/2017 12:17	SMPL
I2151-08	MDABY0	559	0.0971 -		ISM02.4	PPB	3/9/2017 12:19	SMPL
I2151-09	MDABY1	782	0.1435 -		ISM02.4	PPB	3/9/2017 12:21	SMPL
I2151-10	MDABY2	274	0.0377 -		ISM02.4	PPB	3/9/2017 12:23	SMPL
I2151-11	MDABY3	553	0.0958 -		ISM02.4	PPB	3/9/2017 12:26	SMPL
I2151-12	MDABY4	1084	0.2065 -		ISM02.4	PPB	3/9/2017 12:28	SMPL
I2151-13	MDABY5	255	0.0337 -		ISM02.4	PPB	3/9/2017 12:30	SMPL
I2151-14	MDABY6	992	0.1873 -		ISM02.4	PPB	3/9/2017 12:33	SMPL
I2151-15	MDABY7	604	0.1064 -		ISM02.4	PPB	3/9/2017 12:35	SMPL
I2151-16	MDABZ9	297	0.0425 -		ISM02.4	PPB	3/9/2017 12:37	SMPL
I2151-17	MDAC00	1073	0.2042 -		ISM02.4	PPB	3/9/2017 12:39	SMPL
I2151-18	MDAC01	862	0.1602 -		ISM02.4	PPB	3/9/2017 12:42	SMPL
I2151-19	MDAC02	198	0.0218 -		ISM02.4	PPB	3/9/2017 12:44	SMPL
PB97455BL	PBS004	105	0.0025 -		ISM02.4	PPB	3/9/2017 12:46	SMPL

INSTRUMENT ID : CV1

I2153-02	MDABX2	455478	HIGH	-	ISM02.4	PPB	3/9/2017 12:48
I2153-04	MDABX8	-70	-0.034	-	ISM02.4	PPB	3/9/2017 12:50 SMPL
I2153-05	MDABX8D	264	0.0356	-	ISM02.4	PPB	3/9/2017 12:53 SMPL
I2153-06	MDABX8S	12408	2.5663	-	ISM02.4	PPB	3/9/2017 12:55 SMPL
I2153-07	MDABY8	726	0.1319	-	ISM02.4	PPB	3/9/2017 12:57 SMPL
CCV068	CCV068	24409	5.0672	-	ISM02.4	PPB	3/9/2017 12:59 SMPL
CCB068	CCB068	79	-0.003	-	ISM02.4	PPB	3/9/2017 13:02 SMPL
I2153-08	MDABY9	712	0.129	-	ISM02.4	PPB	3/9/2017 13:04 SMPL
I2153-09	MDABZ0	163	0.0145	-	ISM02.4	PPB	3/9/2017 13:06 SMPL
I2153-10	MDABZ1	496	0.0839	-	ISM02.4	PPB	3/9/2017 13:09 SMPL
I2153-11	MDABZ2	889	0.1658	-	ISM02.4	PPB	3/9/2017 13:11 SMPL
I2153-12	MDABZ3	286	0.0402	-	ISM02.4	PPB	3/9/2017 13:13 SMPL
I2153-13	MDABZ4	1500	0.2932	-	ISM02.4	PPB	3/9/2017 13:15 SMPL
I2153-14	MDABZ5	1071	0.2038	-	ISM02.4	PPB	3/9/2017 13:18 SMPL
I2153-15	MDABZ6	121	0.0058	-	ISM02.4	PPB	3/9/2017 13:20 SMPL
I2153-16	MDABZ7	560	0.0973	-	ISM02.4	PPB	3/9/2017 13:22 SMPL
I2153-17	MDABZ8	1628	0.3198	-	ISM02.4	PPB	3/9/2017 13:25 SMPL
I2104-01DLX5	MCOAJ8	30678	6.3736	-	ISM02.4	PPB	3/9/2017 13:27 SMPL
I2153-02DLX10	MDABX2	69050	14.3699	-	ISM02.4	PPB	3/9/2017 13:29 SMPL
I2151-02	MDABX1	12072	2.4963	-	ISM02.4	PPB	3/9/2017 13:34 SMPL
CCV069	CCV069	24337	5.0522	-	ISM02.4	PPB	3/9/2017 13:44 SMPL
CCB069	CCB069	54	-0.0082	-	ISM02.4	PPB	3/9/2017 13:47 SMPL
I2153-02DL2X50	MDABX2	13924	2.8822	-	ISM02.4	PPB	3/9/2017 13:51 SMPL
CCV070	CCV070	24160	5.0153	-	ISM02.4	PPB	3/9/2017 13:58 SMPL
CCB070	CCB070	62	-0.0065	-	ISM02.4	PPB	3/9/2017 14:00 SMPL