

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: _____

ANALYST: JIGNESH

Date: 03/09/2017

Run Number: LB86264

BalanceID: WC SC-3

DATE & TIME IN: 03/09/2017 14:30

DATE & TIME OUT: 03/09/2017 18:35

OVEN TEMP: 104 °C

OVEN TEMP: 103 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86264BL	LB86264BL	1.3258	1.3258	100	1.3258	1.3258	1.3258	0	0
2	LB86264BS	LB86264BS	1.4743	1.4743	100	1.533	1.533	1.533	0.0587	587
3	I2203-01	SPDES-1	1.4272	1.4272	400	1.4273	1.4273	1.4273	0.0001	0.3
4	I2203-01DUP	SPDES-1DUP	1.4049	1.4049	400	1.4051	1.4051	1.4051	0.0002	0.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

2986264

WorkList Name : tss i2203

WorkList ID : 96382

Date : 3/9/2017 1:54:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2203-01	TSS	Cool 4 deg C	SKAN06	L13	SPDES-1	03/09/2017	SM2540 D

Date/Time 03/09/17 1:54 PM
 Received by: J1
 Relinquished by: [Signature]

Date/Time 03/09/17 3:35 PM
 Received by: [Signature]
 Relinquished by: J1