


ISM02.4

LB86280
INSTRUMENT ID: CV2 μ Abs.:

4277E



Linear

A= 0.0000e+000

B= 2.3524e-004

C= -6.2977e-002

Rho= 0.9999794

Accept=Accepted

slope
= 1/23524

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.036	-0.036	116	0.000	116					1
0.2	0.200	0.208	0.008	1154	0.0 %	1154					4
2.5	2.500	2.531	0.031	11026	0.0 %	11026					1
5.0	5.000	5.022	0.022	21615	0.0 %	21615					0
7.5	7.500	7.475	-0.025	32046	0.0 %	32046					0
10.0	10.000	9.999	-0.001	42776	0.0 %	42776					0

NS
3/9/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	116	-		0 ISM02.4	PPB	3/9/2017 19:11	Std
	0.2 S0.2	1154	-		0.2 ISM02.4	PPB	3/9/2017 19:13	Std
	2.5 S2.5	11026	-		2.5 ISM02.4	PPB	3/9/2017 19:16	Std
	5 S5	21615	-		5 ISM02.4	PPB	3/9/2017 19:18	Std
	7.5 S7.5	32046	-		7.5 ISM02.4	PPB	3/9/2017 19:20	Std
	10 S10	42776	-		10 ISM02.4	PPB	3/9/2017 19:23	Std
ICV	ICV	16891	3.9104 -		ISM02.4	PPB	3/9/2017 19:26	SMPL
ICB	ICB	36	-0.0545 -		ISM02.4	PPB	3/9/2017 19:28	SMPL
CCV074	CCV074	21311	4.9501 -		ISM02.4	PPB	3/9/2017 19:30	SMPL
CCB074	CCB074	25	-0.0571 -		ISM02.4	PPB	3/9/2017 19:33	SMPL
PB97481BL	PBW003	349	0.0191 -		ISM02.4	PPB	3/9/2017 19:35	SMPL
I2084-07	MX2879	4116	0.9053 -		ISM02.4	PPB	3/9/2017 19:37	SMPL
I2084-08	MX2879D	4229	0.9318 -		ISM02.4	PPB	3/9/2017 19:39	SMPL
I2084-09	MX2879S	8149	1.854 -		ISM02.4	PPB	3/9/2017 19:42	SMPL
PB97482BL	PBW003	55	-0.05 -		ISM02.4	PPB	3/9/2017 19:44	SMPL
I2093-01	MC0AD9	32	-0.0554 -		ISM02.4	PPB	3/9/2017 19:46	SMPL
I2093-02	MC0AE0	42	-0.0531 -		ISM02.4	PPB	3/9/2017 19:49	SMPL
I2093-03	MC0AE0D	27	-0.0566 -		ISM02.4	PPB	3/9/2017 19:51	SMPL
I2093-04	MC0AE0S	4373	0.9657 -		ISM02.4	PPB	3/9/2017 19:53	SMPL
I2093-05	MC0AE1	-2	-0.0634 -		ISM02.4	PPB	3/9/2017 19:55	SMPL
I2093-06	MC0AE2	22	-0.0578 -		ISM02.4	PPB	3/9/2017 19:58	SMPL
I2093-07	MC0AE3	25	-0.0571 -		ISM02.4	PPB	3/9/2017 20:00	SMPL
I2093-08	MC0AE4	18	-0.0587 -		ISM02.4	PPB	3/9/2017 20:02	SMPL
I2093-09	MC0AE6	26	-0.0569 -		ISM02.4	PPB	3/9/2017 20:04	SMPL
I2093-10	MC0AE7	15	-0.0594 -		ISM02.4	PPB	3/9/2017 20:07	SMPL
I2093-11	MC0AE8	22	-0.0578 -		ISM02.4	PPB	3/9/2017 20:09	SMPL
I2093-12	MC0AE9	18	-0.0587 -		ISM02.4	PPB	3/9/2017 20:11	SMPL
I2093-13	MC0AL4	85	-0.043 -		ISM02.4	PPB	3/9/2017 20:14	SMPL
PB97483BL	PBW003	9	-0.0609 -		ISM02.4	PPB	3/9/2017 20:16	SMPL
I2098-01	MC0AD9	19	-0.0585 -		ISM02.4	PPB	3/9/2017 20:18	SMPL
CCV075	CCV075	21234	4.932 -		ISM02.4	PPB	3/9/2017 20:20	SMPL
CCB075	CCB075	24	-0.0573 -		ISM02.4	PPB	3/9/2017 20:23	SMPL
I2098-02	MC0AE0	-2	-0.0634 -		ISM02.4	PPB	3/9/2017 20:25	SMPL
I2098-03	MC0AE0D	-3	-0.0637 -		ISM02.4	PPB	3/9/2017 20:27	SMPL
I2098-04	MC0AE0S	4210	0.9274 -		ISM02.4	PPB	3/9/2017 20:29	SMPL
I2098-05	MC0AE1	-23	-0.0684 -		ISM02.4	PPB	3/9/2017 20:32	SMPL
I2098-06	MC0AE2	0	-0.063 -		ISM02.4	PPB	3/9/2017 20:34	SMPL
I2098-07	MC0AE3	22	-0.0578 -		ISM02.4	PPB	3/9/2017 20:36	SMPL
I2098-08	MC0AF1	11	-0.0604 -		ISM02.4	PPB	3/9/2017 20:39	SMPL
CCV076	CCV076	21175	4.9181 -		ISM02.4	PPB	3/9/2017 20:43	SMPL
CCB076	CCB076	23	-0.0576 -		ISM02.4	PPB	3/9/2017 20:45	SMPL