

7470A


LB86286

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35493



A= 0.0000e+000

B= 2.8195e-004

C= -7.2554e-002

Rho= 0.9998812

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	225	0.000	225				
0.2	0.200	0.206	0.006	988	0.0 %	988				
2.5	2.500	2.498	-0.002	9117	0.0 %	9117				
5.0	5.000	4.956	-0.044	17836	0.0 %	17836				
7.5	7.500	7.614	0.114	27264	0.0 %	27264				
10.0	10.000	9.935	-0.065	35493	0.0 %	35493				

LB86286 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Mar 2017 10:43:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	225	-	0.0000	7470A	PPB	10 Mar 2017 10:48:33	S	Std
0.2 - 1	50.2	988	-	0.2000	7470A	PPB	10 Mar 2017 10:50:41	S	Std
2.5 - 1	52.5	9117	-	2.5000	7470A	PPB	10 Mar 2017 10:52:49	S	Std
5.0 - 1	55	17836	-	5.0000	7470A	PPB	10 Mar 2017 10:54:58	S	Std
7.5 - 1	57.5	27264	-	7.5000	7470A	PPB	10 Mar 2017 11:02:02	S	Std
10.0 - 1	510	35493	-	10.0000	7470A	PPB	10 Mar 2017 11:06:36	S	Std
ICV65 - 1	ICV65	14307	3.9612	-	7470A	PPB	10 Mar 2017 11:09:17	U	SMPL
ICB65 - 1	ICB65	150	-0.0303	-	7470A	PPB	10 Mar 2017 11:11:23	U	SMPL
CCV77 - 1	CCV77	17652	4.9043	-	7470A	PPB	10 Mar 2017 11:13:30	U	SMPL
CCB77 - 1	CCB77	148	-0.0308	-	7470A	PPB	10 Mar 2017 11:15:36	U	SMPL
CRA - 1	CRA	988	0.2060	-	7470A	PPB	10 Mar 2017 11:17:42	U	SMPL
HighStd - 1	HighStd	35730	10.0014	-	7470A	PPB	10 Mar 2017 11:19:49	U	SMPL
ChkStd - 1	ChkStd	25893	7.2279	-	7470A	PPB	10 Mar 2017 11:21:56	U	SMPL
PB97509BL - 1	PBW	229	-0.0080	-	7470A	PPB	10 Mar 2017 11:24:04	U	SMPL
PB97509BS - 1	LCWS	12992	3.5905	-	7470A	PPB	10 Mar 2017 11:26:12	U	SMPL
PB97509TB - 1	PB97445TB	250	-0.0021	-	7470A	PPB	10 Mar 2017 11:28:19	U	SMPL
I2160-02 - 1	WC-B29-B35-001	641	0.1082	-	7470A	PPB	10 Mar 2017 11:30:28	U	SMPL
I2160-04 - 1	WC-B44-B45-001	530	0.0769	-	7470A	PPB	10 Mar 2017 11:32:36	U	SMPL
I2160-06 - 1	WC-B37-B40-001	302	0.0126	-	7470A	PPB	10 Mar 2017 11:34:44	U	SMPL
I2160-08 - 1	WC-B41-B52-001	350	0.0261	-	7470A	PPB	10 Mar 2017 11:36:51	U	SMPL
CCV78 - 1	CCV78	17454	4.8485	-	7470A	PPB	10 Mar 2017 11:38:58	U	SMPL
CCB78 - 1	CCB78	150	-0.0303	-	7470A	PPB	10 Mar 2017 11:41:05	U	SMPL
I2160-10 - 1	WC-B54-B56-001	2728	0.6966	-	7470A	PPB	10 Mar 2017 11:43:12	U	SMPL
I2172-21 - 1	D-38	363	0.0298	-	7470A	PPB	10 Mar 2017 11:45:19	U	SMPL
I2172-22 - 1	D-41	378	0.0340	-	7470A	PPB	10 Mar 2017 11:47:26	U	SMPL
I2172-23 - 1	D-44	383	0.0354	-	7470A	PPB	10 Mar 2017 11:49:33	U	SMPL
I2172-24 - 1	D-47	538	0.0791	-	7470A	PPB	10 Mar 2017 11:51:40	U	SMPL
I2172-25 - 1	D-49	414	0.0442	-	7470A	PPB	10 Mar 2017 11:53:48	U	SMPL
I2172-26 - 1	D-51	424	0.0470	-	7470A	PPB	10 Mar 2017 11:55:56	U	SMPL
I2172-27 - 1	D-53	354	0.0273	-	7470A	PPB	10 Mar 2017 11:58:04	U	SMPL
I2172-28 - 1	D-55	322	0.0182	-	7470A	PPB	10 Mar 2017 12:00:13	U	SMPL
I2172-29 - 1	D-57	407	0.0422	-	7470A	PPB	10 Mar 2017 12:02:20	U	SMPL
CCV79 - 1	CCV79	17741	4.9294	-	7470A	PPB	10 Mar 2017 12:04:27	U	SMPL
CCB79 - 1	CCB79	179	-0.0221	-	7470A	PPB	10 Mar 2017 12:06:34	U	SMPL
I2172-30 - 1	D-61	384	0.0357	-	7470A	PPB	10 Mar 2017 12:08:41	U	SMPL
I2172-30DUP - 1	D-61DUP	370	0.0318	-	7470A	PPB	10 Mar 2017 12:10:48	U	SMPL
I2172-30MS - 1	D-61MS	13038	3.6035	-	7470A	PPB	10 Mar 2017 12:12:55	U	SMPL
I2172-30MSD - 1	D-61MSD	13088	3.6175	-	7470A	PPB	10 Mar 2017 12:15:02	U	SMPL
I2172-30LX5 - 1		208	-0.0139	-	7470A	PPB	10 Mar 2017 12:17:10	U	SMPL
I2172-30A - 1		13786	3.8143	-	7470A	PPB	10 Mar 2017 12:19:17	U	SMPL
CCV80 - 1	CCV80	17736	4.9280	-	7470A	PPB	10 Mar 2017 12:21:26	U	SMPL
CCB80 - 1	CCB80	171	-0.0243	-	7470A	PPB	10 Mar 2017 12:23:34	U	SMPL