

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 03/11/2017

Run Number: LB86288

DATE & TIME IN: 03/10/2017 17:00

OVEN TEMP: 98 °C

BalanceID: WC SC-3

DATE & TIME OUT: 03/11/2017 12:50

OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @105°C (g)	2nd Dish+Sample weight after Drying @105°C (g)	Final Dish+Sample weight after Drying @105°C (g)	Weight (g)	Result %
1	LB86288BL	LB86288BL	65.0237	65.0237	100	65.0237	65.0237	65.0237	0	0
2	I2128-01	D-1	63.828	63.828	100	63.8294	63.8294	63.8294	0.0014	14
3	I2128-01DUP	D-1DUP	64.1196	64.1196	100	64.121	64.121	64.121	0.0014	14
4	I2128-02	D-8	89.2187	89.2187	100	89.2488	89.2488	89.2488	0.0301	301
5	I2128-03	D-12	79.5116	79.5116	100	79.5184	79.5184	79.5184	0.0068	68
6	I2128-04	D-14	75.078	75.078	100	75.0814	75.0814	75.0814	0.0034	34
7	I2128-05	D-18	68.8505	68.8505	100	68.8551	68.8551	68.8551	0.0046	46
8	I2128-06	D-21	81.5109	81.5109	100	81.517	81.517	81.517	0.0061	61
9	I2128-07	D-23	75.5648	75.5648	100	75.57	75.57	75.57	0.0052	52
10	I2128-08	D-25	79.5475	79.5475	100	79.5527	79.5527	79.5527	0.0052	52
11	I2128-09	D-26	74.85	74.85	100	74.8536	74.8536	74.8536	0.0036	36
12	I2130-01	D-27	151.231	151.231	100	151.231	151.231	151.231	0	0
13	I2130-02	D-30	152.979	152.979	100	152.979	152.979	152.979	0	0
14	I2130-03	D-33	143.61	143.61	100	143.612	143.612	143.612	0.002	20
15	I2130-04	D-35	149.259	149.259	100	149.261	149.261	149.261	0.002	20

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @105°C (g)	2nd Dish+Sample weight after Drying @105°C (g)	Final Dish+Sample weight after Drying @105°C (g)	Weight (g)	Result %
A = Sample Volume (ml) B = Empty Dish Weight (g) C = Final Empty Dish + Sample weight after drying @105°C(g) D = Weight (g)										
						Weight (g) = C - B				
						Result mg/L = $\frac{D}{A} \times 1000$				

LB 86288

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS I2027 WorkList ID : 96412 Date : 3/10/2017 11:30:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2128-01	ASTM TS	Cool 4 deg C	SEVE02	K43	D-1	03/06/2017	SM2540 B
	Solid	I2128-02	ASTM TS	Cool 4 deg C	SEVE02	K43	D-8	03/06/2017	SM2540 B
	Solid	I2128-03	ASTM TS	Cool 4 deg C	SEVE02	K43	D-12	03/06/2017	SM2540 B
	Solid	I2128-04	ASTM TS	Cool 4 deg C	SEVE02	K43	D-14	03/06/2017	SM2540 B
	Solid	I2128-05	ASTM TS	Cool 4 deg C	SEVE02	K43	D-18	03/06/2017	SM2540 B
	Solid	I2128-06	ASTM TS	Cool 4 deg C	SEVE02	K43	D-21	03/06/2017	SM2540 B
	Solid	I2128-07	ASTM TS	Cool 4 deg C	SEVE02	K43	D-23	03/06/2017	SM2540 B
	Solid	I2128-08	ASTM TS	Cool 4 deg C	SEVE02	K43	D-25	03/06/2017	SM2540 B
	Solid	I2128-09	ASTM TS	Cool 4 deg C	SEVE02	K43	D-26	03/06/2017	SM2540 B
	Solid	I2130-01	ASTM TS	Cool 4 deg C	SEVE02	K43	D-27	03/06/2017	SM2540 B
	Solid	I2130-02	ASTM TS	Cool 4 deg C	SEVE02	K43	D-30	03/06/2017	SM2540 B
	Solid	I2130-03	ASTM TS	Cool 4 deg C	SEVE02	K43	D-33	03/06/2017	SM2540 B
	Solid	I2130-04	ASTM TS	Cool 4 deg C	SEVE02	K43	D-35	03/06/2017	SM2540 B

Date/Time 03/10/17 11:30 AM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 03/10/17 6:35 PM
 Received by: [Signature]
 Relinquished by: [Signature]