

245.1


LB 86292

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3334E



A= 0.0000e+000

B= 3.0017e-004

C= -9.4564e-002

Rho= 0.9998684

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	328	0.000	328				
0.2	0.200	0.174	-0.026	896	0.0 %	896				
2.5	2.500	2.495	-0.005	8628	0.0 %	8628				
5.0	5.000	4.996	-0.004	16958	0.0 %	16958				
7.5	7.500	7.616	0.116	25688	0.0 %	25688				
10.0	10.000	9.915	-0.085	33345	0.0 %	33345				

LB86292 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Mar 2017 14:30:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	328	-	0.0000	245.1	PPB	10 Mar 2017 14:35:09	S	Std
0.2 - 1	30.2	896	-	0.2000	245.1	PPB	10 Mar 2017 14:37:17	S	Std
2.5 - 1	52.5	8628	-	2.5000	245.1	PPB	10 Mar 2017 14:39:26	S	Std
5.0 - 1	55	16958	-	5.0000	245.1	PPB	10 Mar 2017 14:41:34	S	Std
7.5 - 1	57.5	25688	-	7.5000	245.1	PPB	10 Mar 2017 14:43:41	S	Std
10.0 - 1	510	33345	-	10.0000	245.1	PPB	10 Mar 2017 14:45:49	S	Std
ICV66 - 1	ICV66	13544	3.9709	-245.1		PPB	10 Mar 2017 14:48:44	U	SMPL
ICB66 - 1	ICB66	318	0.0009	-245.1		PPB	10 Mar 2017 14:50:50	U	SMPL
CCV81 - 1	CCV81	16885	4.9738	-245.1		PPB	10 Mar 2017 14:58:49	U	SMPL
CCB81 - 1	CCB81	237	-0.0234	-245.1		PPB	10 Mar 2017 15:00:55	U	SMPL
CRA - 1	CRA	929	0.1843	-245.1		PPB	10 Mar 2017 15:03:02	U	SMPL
HighStd - 1	HighStd	33667	10.0112	-245.1		PPB	10 Mar 2017 15:05:08	U	SMPL
ChkStd - 1	ChkStd	23787	7.0455	-245.1		PPB	10 Mar 2017 15:07:15	U	SMPL
PB97507BL - 1	PBW	234	-0.0243	-245.1		PPB	10 Mar 2017 15:09:23	U	SMPL
PB97507BS - 1	LCSW	12308	3.5999	-245.1		PPB	10 Mar 2017 15:11:31	U	SMPL
I2190-01 - 1	030617-DSDC-E-D-M	470	0.0465	-245.1		PPB	10 Mar 2017 15:21:06	U	SMPL
I2190-02 - 1	030617-DSDC-E-U-M	1383	0.3206	-245.1		PPB	10 Mar 2017 15:27:51	U	SMPL
I2190-02DUP - 1	030617-DSDC-E-U-MDUP	1353	0.3116	-245.1		PPB	10 Mar 2017 15:29:59	U	SMPL
I2190-02MS - 1	030617-DSDC-E-U-MMS	13858	4.0652	-245.1		PPB	10 Mar 2017 15:32:07	U	SMPL
I2190-02MSD - 1	030617-DSDC-E-U-MMSD	13814	4.0520	-245.1		PPB	10 Mar 2017 15:34:13	U	SMPL
CCV82 - 1	CCV82	16915	4.9828	-245.1		PPB	10 Mar 2017 15:36:20	U	SMPL
CCB82 - 1	CCB82	154	-0.0483	-245.1		PPB	10 Mar 2017 15:38:27	U	SMPL
I2190-02LX5 - 1		600	0.0855	-245.1		PPB	10 Mar 2017 15:40:34	U	SMPL
I2190-02A - 1		14659	4.3056	-245.1		PPB	10 Mar 2017 15:42:41	U	SMPL
CCV83 - 1	CCV83	17079	5.0320	-245.1		PPB	10 Mar 2017 15:44:48	U	SMPL
CCB83 - 1	CCB83	150	-0.0495	-245.1		PPB	10 Mar 2017 15:46:55	U	SMPL