



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86296

BalanceID: NA

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP53429
COD CCV std, 50ppm	WP53430

Temp In (C): <u>150</u>	Date In: <u>03/10/2017</u>	Time In: <u>11:50</u>
Temp Out (C): <u>150</u>	Date Out: <u>03/10/2017</u>	Time Out: <u>13:50</u>

Intercept: -0.7821

Slope: 1.0223

Regression: 0.9998

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/20/2017	14:05
2	CAL2	10	1	Water	9.000	02/20/2017	14:05
3	CAL3	50	1	Water	49.000	02/20/2017	14:06
4	CAL4	100	1	Water	103.000	02/20/2017	14:06
5	CAL5	150	1	Water	152.000	02/20/2017	14:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86296

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	48.696	02/20/2017	14:07
2	ICB				1	Water	0.000	0.765	02/20/2017	14:07
3	CCV1	50			1	Water	53.000	52.609	03/10/2017	14:51
4	CCB1				1	Water	0.000	0.765	03/10/2017	14:51
5	LB86296BL		50	50	1	Water	0.000	0.765	03/10/2017	14:52
6	LB86296BS	50	50	50	1	Water	48.000	47.718	03/10/2017	14:52
7	I2128-01		50	50	1	Water	18.000	18.372	03/10/2017	14:53
8	I2128-01DUP		50	50	1	Water	17.000	17.394	03/10/2017	14:53
9	I2128-01MS	50	50	50	1	Water	69.000	68.260	03/10/2017	14:54
10	I2128-01MSD	50	50	50	1	Water	68.000	67.282	03/10/2017	14:54
11	I2128-02		50	50	1	Water	150.000	147.493	03/10/2017	14:55
12	I2128-03		50	50	1	Water	83.000	81.955	03/10/2017	14:55
13	I2128-04		50	50	1	Water	13.000	13.481	03/10/2017	14:56
14	I2128-05		50	50	1	Water	10.000	10.547	03/10/2017	14:56
15	CCV2	50			1	Water	52.000	51.631	03/10/2017	14:57
16	CCB2				1	Water	0.000	0.765	03/10/2017	14:57
17	I2128-06		50	50	1	Water	15.000	15.438	03/10/2017	14:58
18	I2128-07		50	50	1	Water	24.000	24.242	03/10/2017	14:58
19	I2128-08		50	50	1	Water	33.000	33.045	03/10/2017	14:59
20	I2128-09		50	50	1	Water	20.000	20.329	03/10/2017	14:59
21	I2130-01		50	50	1	Water	8.000	8.591	03/10/2017	15:00
22	I2130-02		50	50	1	Water	19.000	19.351	03/10/2017	15:00
23	I2130-03		50	50	1	Water	15.000	15.438	03/10/2017	15:01
24	I2130-04		50	50	1	Water	9.000	9.569	03/10/2017	15:01
25	CCV3	50			1	Water	52.000	51.631	03/10/2017	15:02
26	CCB3				1	Water	0.000	0.765	03/10/2017	15:02

2886296

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD031017

WorkList ID : 96410

Date : 3/10/2017 10:48:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2128-01	ASTM COD	Cool 4 deg C	SEVE02	K43	D-1	03/06/2017	SM5220 D
	Solid	I2128-02	ASTM COD	Cool 4 deg C	SEVE02	K43	D-8	03/06/2017	SM5220 D
	Solid	I2128-03	ASTM COD	Cool 4 deg C	SEVE02	K43	D-12	03/06/2017	SM5220 D
	Solid	I2128-04	ASTM COD	Cool 4 deg C	SEVE02	K43	D-14	03/06/2017	SM5220 D
	Solid	I2128-05	ASTM COD	Cool 4 deg C	SEVE02	K43	D-18	03/06/2017	SM5220 D
	Solid	I2128-06	ASTM COD	Cool 4 deg C	SEVE02	K43	D-21	03/06/2017	SM5220 D
	Solid	I2128-07	ASTM COD	Cool 4 deg C	SEVE02	K43	D-23	03/06/2017	SM5220 D
	Solid	I2128-08	ASTM COD	Cool 4 deg C	SEVE02	K43	D-25	03/06/2017	SM5220 D
	Solid	I2128-09	ASTM COD	Cool 4 deg C	SEVE02	K43	D-26	03/06/2017	SM5220 D
	Solid	I2130-01	ASTM COD	Cool 4 deg C	SEVE02	K43	D-27	03/06/2017	SM5220 D
	Solid	I2130-02	ASTM COD	Cool 4 deg C	SEVE02	K43	D-30	03/06/2017	SM5220 D
	Solid	I2130-03	ASTM COD	Cool 4 deg C	SEVE02	K43	D-33	03/06/2017	SM5220 D
	Solid	I2130-04	ASTM COD	Cool 4 deg C	SEVE02	K43	D-35	03/06/2017	SM5220 D

Date/Time 03/10/17 11:00

Received by: IB

Relinquished by: IB

Date/Time 03/10/17 13:48

Received by: IB

Relinquished by: IB