


7471A

LB 86299
INSTRUMENT ID: CVL

Linear

 μ Abs.:

50720



A= 0.0000e+000

B= 1.9804e-004

C= 2.6380e-002

Rho= 0.9998051

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.054	0.054	141	0.000	141				
0.2	0.200	0.204	0.004	897	0.0 %	897				
2.5	2.500	2.422	-0.078	12096	0.0 %	12096				
5.0	5.000	5.065	0.065	25444	0.0 %	25444				
7.5	7.500	7.384	-0.116	37152	0.0 %	37152				
10.0	10.000	10.071	0.071	50720	0.0 %	50720				

LB86299 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 10 Mar 2017 16:41:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.0000	7471A	PPB	10 Mar 2017 17:02:05	S	Std
0.2 - 1	50-2	897	-	0.2000	7471A	PPB	10 Mar 2017 17:04:22	S	Std
2.5 - 1	52.5	12096	-	2.5000	7471A	PPB	10 Mar 2017 17:06:40	S	Std
5.0 - 1	55	25444	-	5.0000	7471A	PPB	10 Mar 2017 17:08:57	S	Std
7.5 - 1	57.5	37152	-	7.5000	7471A	PPB	10 Mar 2017 17:13:53	S	Std
10.0 - 1	510	50720	-	10.0000	7471A	PPB	10 Mar 2017 17:18:45	S	Std
ICV67 - 1	ICV67	18314	3.6532	-	7471A	PPB	10 Mar 2017 17:21:33	U	SMPL
ICB67 - 1	ICB67	84	0.0430	-	7471A	PPB	10 Mar 2017 17:23:48	U	SMPL
CCV84 - 1	CCV84	24835	4.9446	-	7471A	PPB	10 Mar 2017 17:28:37	U	SMPL
CCB84 - 1	CCB84	77	0.0416	-	7471A	PPB	10 Mar 2017 17:30:52	U	SMPL
CRA - 1	CRA	898	0.2042	-	7471A	PPB	10 Mar 2017 17:33:08	U	SMPL
HighStd - 1	HighStd	51109	10.1478	-	7471A	PPB	10 Mar 2017 17:35:24	U	SMPL
ChkStd - 1	ChkStd	34837	6.9254	-	7471A	PPB	10 Mar 2017 17:41:05	U	SMPL
PB97508BL - 1	PBS	66	0.0395	-	7471A	PPB	10 Mar 2017 17:43:21	U	SMPL
PB97508BS - 1	LCSS	18338	3.6580	-	7471A	PPB	10 Mar 2017 17:45:38	U	SMPL
I2195-01 - 1	TP-9-COMP-0-12FT	1074	0.2391	-	7471A	PPB	10 Mar 2017 17:55:42	U	SMPL
I2195-01DUP - 1	TP-9-COMP-0-12FTDUP	1057	0.2357	-	7471A	PPB	10 Mar 2017 17:57:58	U	SMPL
I2195-01MS - 1	TP-9-COMP-0-12FTMS	17604	3.5126	-	7471A	PPB	10 Mar 2017 18:00:16	U	SMPL
I2195-01MSD - 1	TP-9-COMP-0-12FTMSD	17417	3.4756	-	7471A	PPB	10 Mar 2017 18:02:34	U	SMPL
I2195-06 - 1	TP-7-TP-6-COMP-0-12FT	1032	0.2308	-	7471A	PPB	10 Mar 2017 18:04:49	U	SMPL
CCV85 - 1	CCV85	24977	4.9727	-	7471A	PPB	10 Mar 2017 18:07:06	U	SMPL
CCB85 - 1	CCB85	113	0.0488	-	7471A	PPB	10 Mar 2017 18:09:22	U	SMPL
I2195-11 - 1	TP-5-TP-4-COMP-0-12FT	1141	0.2523	-	7471A	PPB	10 Mar 2017 18:11:38	U	SMPL
I2207-02 - 1	41D-9-11	356	0.0969	-	7471A	PPB	10 Mar 2017 18:13:54	U	SMPL
I2209-01 - 1	DB-1(0.0-0.5)	6622	1.3378	-	7471A	PPB	10 Mar 2017 18:16:10	U	SMPL
I2209-02 - 1	DB-2(0.0-0.5)	2570	0.5353	-	7471A	PPB	10 Mar 2017 18:18:27	U	SMPL
I2209-03 - 1	DB-4(0.0-0.5)	3367	0.6932	-	7471A	PPB	10 Mar 2017 18:20:43	U	SMPL
I2209-04 - 1	DB-3(0.0-0.5)	2679	0.5569	-	7471A	PPB	10 Mar 2017 18:23:00	U	SMPL
I2209-05 - 1	DB-6A(0.0-0.5)	5212	1.0585	-	7471A	PPB	10 Mar 2017 18:25:17	U	SMPL
I2209-07 - 1	DB-5(0.0-0.5)	5998	1.2142	-	7471A	PPB	10 Mar 2017 18:27:35	U	SMPL
I2209-11 - 1	DUP	2518	0.5250	-	7471A	PPB	10 Mar 2017 18:29:52	U	SMPL
I2211-01 - 1	TP-3-COMP-0-12FT	1124	0.2490	-	7471A	PPB	10 Mar 2017 18:32:09	U	SMPL
CCV86 - 1	CCV86	24891	4.9557	-	7471A	PPB	10 Mar 2017 18:34:25	U	SMPL
CCB86 - 1	CCB86	121	0.0503	-	7471A	PPB	10 Mar 2017 18:36:41	U	SMPL
I2211-06 - 1	TP-2-TP-1-COMP-0-12FT	1167	0.2575	-	7471A	PPB	10 Mar 2017 18:38:57	U	SMPL
I2215-01 - 1	E2-V4-BASE	44786	8.8956	-	7471A	PPB	10 Mar 2017 18:41:14	U	SMPL
I2215-02 - 1	E2-V4-WSW	8504	1.7105	-	7471A	PPB	10 Mar 2017 18:43:30	U	SMPL
I2215-03 - 1	E2-V5-BASE	57711	11.4553	-	7471A	PPB	10 Mar 2017 18:45:46	U	SMPL
I2195-01LX5 - 1		221	0.0701	-	7471A	PPB	10 Mar 2017 18:48:03	U	SMPL
I2195-01A - 1		18861	3.7615	-	7471A	PPB	10 Mar 2017 18:50:20	U	SMPL
CCV87 - 1	CCV87	25464	5.0692	-	7471A	PPB	10 Mar 2017 18:52:37	U	SMPL
CCB87 - 1	CCB87	112	0.0486	-	7471A	PPB	10 Mar 2017 18:54:54	U	SMPL
I2215-03DLX5 - 1		11457	2.2953	-	7471A	PPB	10 Mar 2017 19:14:28	U	SMPL
CCV88 - 1	CCV88	25084	4.9939	-	7471A	PPB	10 Mar 2017 19:16:43	U	SMPL
CCB88 - 1	CCB88	105	0.0472	-	7471A	PPB	10 Mar 2017 19:19:00	U	SMPL