



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86307

Temp.: 19.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	03/10/2017	18:25
2	CAL2	1	Water			7.01	03/10/2017	18:26
3	CAL3	1	Water			10.02	03/10/2017	18:27
4	ICV	1	Water			7.02	03/10/2017	18:28
5	CCV1	1	Water			2.01	03/10/2017	18:30
6	I2242-01	1	Water			7.03	03/10/2017	18:31
7	I2242-01DUP	1	Water			7.04	03/10/2017	18:32
8	CCV2	1	Water			12.01	03/10/2017	18:33

25 86307

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12242

WorkList ID : 96434

Date : 3/10/2017 6:10:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2242-01	pH	Cool 4 deg C	TECT01	L31	TEC-GW	03/10/2017	SM 4500-PH B

Date/Time 03-10-17 6:10 PM
Received by: JS
Relinquished by: JS

Date/Time 03-10-17 7:00 AM
Received by: CP
Relinquished by: JS