



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86325

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/10/2017	17:20
2	CAL2	1	Water			7.02	03/10/2017	17:21
3	CAL3	1	Water			10.02	03/10/2017	17:22
4	ICV	1	Water			7.01	03/10/2017	17:23
5	CCV1	1	Water			2.02	03/10/2017	17:24
6	I2230-02	1	Solid	20.03	20	8.27	03/10/2017	17:25
7	I2230-02DUP	1	Solid	20.04	20	8.28	03/10/2017	17:26
8	I2230-04	1	Solid	20.02	20	8.47	03/10/2017	17:28
9	I2230-06	1	Solid	20.06	20	8.55	03/10/2017	17:30
10	I2230-08	1	Solid	20.03	20	8.42	03/10/2017	17:31
11	I2233-02	1	Solid	20.04	20	10.85	03/10/2017	17:32
12	CCV2	1	Water			12.02	03/10/2017	17:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2230

WorkList ID : 96428

Date : 3/10/2017 4:54:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2230-02	Corrosivity	Cool 4 deg C	ENVI27	L31	COMPOSITE-1-CHAR	03/08/2017	9045C
	Solid	I2230-04	Corrosivity	Cool 4 deg C	ENVI27	L31	COMPOSITE-2-CHAR	03/08/2017	9045C
	Solid	I2230-06	Corrosivity	Cool 4 deg C	ENVI27	L31	COMPOSITE-3-CHAR	03/08/2017	9045C
	Solid	I2230-08	Corrosivity	Cool 4 deg C	ENVI27	L31	COMPOSITE-4-CHAR	03/08/2017	9045C
03/15/2017	Solid	I2233-02	Corrosivity	Cool 4 deg C	PSEG01	K42	FRAC TANK BOTTOMS	03/10/2017	9045C

Date/Time 03-10-17 4:54 PM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 03-10-17 4:54 PM
 Received by: [Signature]
 Relinquished by: [Signature]