



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86330

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/09/2017	17:30
2	CAL2	1	Water			7.01	03/09/2017	17:32
3	CAL3	1	Water			10.04	03/09/2017	17:33
4	ICV	1	Water			6.99	03/09/2017	17:34
5	CCV1	1	Water			2.02	03/09/2017	17:35
6	I2208-02	1	Solid	20.02	20	8.96	03/09/2017	17:36
7	I2208-02DUP	1	Solid	20.03	20	8.97	03/09/2017	17:37
8	I2208-04	1	Solid	20.04	20	9.44	03/09/2017	17:38
9	I2211-05	1	Solid	20.03	20	5.47	03/09/2017	17:39
10	I2211-10	1	Solid	20.05	20	5.66	03/09/2017	17:40
11	CCV2	1	Water			12.02	03/09/2017	17:41

WORKLIST(Hardcopy Internal Chain)

LB 86336

WorkList Name : CORROSIONITY I2208

WorkList ID : 96394

Date : 3/9/2017 5:21:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2208-02	Corrosivity	Cool 4 deg C	VIAS01	K12	WC-AGG1-08	03/08/2017	9045C
	Solid	I2208-04	Corrosivity	Cool 4 deg C	VIAS01	K12	WC-AGG2-09	03/08/2017	9045C
03/14/2017	Solid	I2211-05	Corrosivity	Cool 4 deg C	PSEG01	K33	TP-3-COMP-0-12FT	03/09/2017	9045C
03/14/2017	Solid	I2211-10	Corrosivity	Cool 4 deg C	PSEG01	K33	TP-2-TP-1-COMP-0-12FT	03/09/2017	9045C

Date/Time 03-09-17 5:21 PM
 Received by: J11
 Relinquished by: AS

Date/Time 03-09-17 7:00 PM
 Received by: CP
 Relinquished by: SH