

7470A


LB 86334

INSTRUMENT ID: CY1

Linear

 μ Abs.:

34689



A= 0.0000e+000

B= 2.8440e-004

C= -5.6355e-003

Rho= 0.9995717

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	47	0.000	47				
0.2	0.200	0.216	0.016	778	0.0 %	778				
2.5	2.500	2.375	-0.125	8371	0.0 %	8371				
5.0	5.000	5.069	0.069	17843	0.0 %	17843				
7.5	7.500	7.672	0.172	26997	0.0 %	26997				
10.0	10.000	9.860	-0.140	34689	0.0 %	34689				

LB86334 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Mar 2017 10:47:59

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	47	-	0.0000	7470A	PPB	13 Mar 2017 11:18:46	S	Std
0.2 - 1	50.2	778	-	0.2000	7470A	PPB	13 Mar 2017 11:20:55	S	Std
2.5 - 1	52.5	8371	-	2.5000	7470A	PPB	13 Mar 2017 11:23:03	S	Std
5.0 - 1	85	17843	-	5.0000	7470A	PPB	13 Mar 2017 11:25:11	S	Std
7.5 - 1	57.5	26997	-	7.5000	7470A	PPB	13 Mar 2017 11:30:15	S	Std
10.0 - 1	57.0	34689	-	10.0000	7470A	PPB	13 Mar 2017 11:34:58	S	Std
ICV68 - 1	ICV68	14337	4.0719	-	7470A	PPB	13 Mar 2017 11:37:40	U	SMPL
ICB68 - 1	ICB68	13	-0.0019	-	7470A	PPB	13 Mar 2017 11:39:46	U	SMPL
CCV89 - 1	CCV89	17917	5.0900	-	7470A	PPB	13 Mar 2017 11:41:53	U	SMPL
CCB89 - 1	CCB89	-12	-0.0090	-	7470A	PPB	13 Mar 2017 11:43:59	U	SMPL
CRA - 1	CRA	784	0.2173	-	7470A	PPB	13 Mar 2017 11:46:06	U	SMPL
HighStd - 1	HighStd	35386	10.0583	-	7470A	PPB	13 Mar 2017 11:48:12	U	SMPL
ChkStd - 1	ChkStd	24784	7.0431	-	7470A	PPB	13 Mar 2017 11:50:19	U	SMPL
PB97522BL - 1	PBW	0	-0.0056	-	7470A	PPB	13 Mar 2017 11:52:27	U	SMPL
PB97522BS - 1	LCSW	14724	4.1819	-	7470A	PPB	13 Mar 2017 11:54:34	U	SMPL
PB97522TB - 1	PB97495TB	23	0.0009	-	7470A	PPB	13 Mar 2017 11:56:42	U	SMPL
I2174-19 - 1	D-64	128	0.0308	-	7470A	PPB	13 Mar 2017 11:58:50	U	SMPL
I2174-19DUP - 1	D-64DUP	144	0.0353	-	7470A	PPB	13 Mar 2017 12:03:27	U	SMPL
I2174-19MS - 1	D-64MS	14060	3.9931	-	7470A	PPB	13 Mar 2017 12:05:35	U	SMPL
I2174-19MSD - 1	D-64MSD	14432	4.0989	-	7470A	PPB	13 Mar 2017 12:07:42	U	SMPL
CCV90 - 1	CCV90	17931	5.0940	-	7470A	PPB	13 Mar 2017 12:09:49	U	SMPL
CCB90 - 1	CCB90	-3	-0.0065	-	7470A	PPB	13 Mar 2017 12:11:55	U	SMPL
I2174-20 - 1	D-67	199	0.0510	-	7470A	PPB	13 Mar 2017 12:14:02	U	SMPL
I2174-21 - 1	D-69	607	0.1670	-	7470A	PPB	13 Mar 2017 12:16:09	U	SMPL
I2174-22 - 1	D-71	271	0.0714	-	7470A	PPB	13 Mar 2017 12:18:16	U	SMPL
I2174-23 - 1	D-74	332	0.0888	-	7470A	PPB	13 Mar 2017 12:20:23	U	SMPL
I2174-24 - 1	D-77	294	0.0780	-	7470A	PPB	13 Mar 2017 12:22:31	U	SMPL
I2174-25 - 1	D-79	205	0.0527	-	7470A	PPB	13 Mar 2017 12:24:38	U	SMPL
I2174-26 - 1	D-81	123	0.0293	-	7470A	PPB	13 Mar 2017 12:26:47	U	SMPL
I2174-27 - 1	D-85	106	0.0245	-	7470A	PPB	13 Mar 2017 12:28:55	U	SMPL
I2195-05 - 1	TP-9-COMP-0-12FT	87	0.0191	-	7470A	PPB	13 Mar 2017 12:31:03	U	SMPL
I2195-10 - 1	TP-7-TP-6-COMP-0-12FT	111	0.0259	-	7470A	PPB	13 Mar 2017 12:33:10	U	SMPL
CCV91 - 1	CCV91	17822	5.0630	-	7470A	PPB	13 Mar 2017 12:35:17	U	SMPL
CCB91 - 1	CCB91	-12	-0.0090	-	7470A	PPB	13 Mar 2017 12:37:24	U	SMPL
I2195-15 - 1	TP-5-TP-4-COMP-0-12FT	106	0.0245	-	7470A	PPB	13 Mar 2017 12:39:31	U	SMPL
I2208-02 - 1	WC-AGG1-08	154	0.0382	-	7470A	PPB	13 Mar 2017 12:41:38	U	SMPL
I2208-04 - 1	WC-AGG2-09	114	0.0268	-	7470A	PPB	13 Mar 2017 12:43:44	U	SMPL
I2211-05 - 1	TP-3-COMP-0-12FT	37	0.0049	-	7470A	PPB	13 Mar 2017 12:45:51	U	SMPL
I2211-10 - 1	TP-2-TP-1-COMP-0-12FT	59	0.0111	-	7470A	PPB	13 Mar 2017 12:47:59	U	SMPL
I2174-19LX5 - 1		77	0.0163	-	7470A	PPB	13 Mar 2017 12:50:07	U	SMPL
I2174-19A - 1		14319	4.0668	-	7470A	PPB	13 Mar 2017 12:52:15	U	SMPL
CCV92 - 1	CCV92	17808	5.0590	-	7470A	PPB	13 Mar 2017 12:54:23	U	SMPL
CCB92 - 1	CCB92	-13	-0.0093	-	7470A	PPB	13 Mar 2017 12:56:31	U	SMPL