



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86337

Temp.: 18.2 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/13/2017	13:35
2	CAL2	1	Water			7.00	03/13/2017	13:36
3	CAL3	1	Water			10.03	03/13/2017	13:37
4	ICV	1	Water			7.00	03/13/2017	13:38
5	CCV1	1	Water			2.01	03/13/2017	13:40
6	I2253-01	1	Water			3.94	03/13/2017	13:41
7	I2253-01DUP	1	Water			3.95	03/13/2017	13:42
8	CCV2	1	Water			12.01	03/13/2017	13:43

LBG633+

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12253 WorkList ID : 96482 Date : 3/13/2017 1:24:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2253-01	pH	NONE	STER08	L13	POOLWATER	03/08/2017	SM 4500-PH B

Date/Time 03-13-17 1:24 PM
Received by: SP
Relinquished by: _____

Date/Time 03-13-17 2:35 PM
Received by: SP
Relinquished by: SP