

245.1


LB 86343

INSTRUMENT ID: CV2

Linear

 μ Abs.:

34642



A= 0.0000e+000

B= 2.8953e-004

C= 2.0524e-002

Rho= 0.9999345

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	31	0.000	31				
0.2	0.200	0.222	0.022	695	0.0 %	695				
2.5	2.500	2.492	-0.008	8536	0.0 %	8536				
5.0	5.000	4.919	-0.081	16918	0.0 %	16918				
7.5	7.500	7.487	-0.013	25789	0.0 %	25789				
10.0	10.000	10.051	0.051	34642	0.0 %	34642				

LB86343 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Mar 2017 14:56:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	245.1	PPB	13 Mar 2017 14:59:51	S	Std
0.2 - 1	50.2	695	-	0.2000	245.1	PPB	13 Mar 2017 15:01:59	S	Std
2.5 - 1	52.5	8536	-	2.5000	245.1	PPB	13 Mar 2017 15:04:07	S	Std
5.0 - 1	55	16918	-	5.0000	245.1	PPB	13 Mar 2017 15:06:16	S	Std
7.5 - 1	57.5	25789	-	7.5000	245.1	PPB	13 Mar 2017 15:08:24	S	Std
10.0 - 1	510	34642	-	10.0000	245.1	PPB	13 Mar 2017 15:10:31	S	Std
ICV69 - 1	ICV69	13260	3.8597	-245.1	PPB	PPB	13 Mar 2017 15:13:09	U	SMPL
ICB69 - 1	ICB69	-43	0.0081	-245.1	PPB	PPB	13 Mar 2017 15:15:15	U	SMPL
CCV93 - 1	CCV93	16871	4.9053	-245.1	PPB	PPB	13 Mar 2017 15:17:22	U	SMPL
CCB93 - 1	CCB93	-47	0.0069	-245.1	PPB	PPB	13 Mar 2017 15:19:28	U	SMPL
CRA - 1	CRA	685	0.2189	-245.1	PPB	PPB	13 Mar 2017 15:21:35	U	SMPL
HighStd - 1	HighStd	34456	9.9967	-245.1	PPB	PPB	13 Mar 2017 15:23:42	U	SMPL
ChkStd - 1	ChkStd	23153	6.7241	-245.1	PPB	PPB	13 Mar 2017 15:25:49	U	SMPL
PB97523BL - 1	PBW	55	0.0364	-245.1	PPB	PPB	13 Mar 2017 15:27:56	U	SMPL
PB97523BS - 1	LCSW	13623	3.9648	-245.1	PPB	PPB	13 Mar 2017 15:30:04	U	SMPL
I2190-02 - 1	030617-DSDC-E-U-M	25	0.0278	-245.1	PPB	PPB	13 Mar 2017 15:32:11	U	SMPL
I2224-02 - 1	030917-PRRT-F-D-M	87	0.0457	-245.1	PPB	PPB	13 Mar 2017 15:36:28	U	SMPL
I2231-01 - 1	32220008-3	65	0.0393	-245.1	PPB	PPB	13 Mar 2017 15:38:36	U	SMPL
I2231-01DUP - 1	32220008-3DUP	43	0.0330	-245.1	PPB	PPB	13 Mar 2017 15:40:43	U	SMPL
CCV94 - 1	CCV94	17020	4.9484	-245.1	PPB	PPB	13 Mar 2017 15:42:50	U	SMPL
CCB94 - 1	CCB94	-40	0.0089	-245.1	PPB	PPB	13 Mar 2017 15:44:56	U	SMPL
I2231-01MS - 1	32220008-3MS	12965	3.7743	-245.1	PPB	PPB	13 Mar 2017 15:47:03	U	SMPL
I2231-01MSD - 1	32220008-3MSD	13362	3.8893	-245.1	PPB	PPB	13 Mar 2017 15:49:10	U	SMPL
I2232-01 - 1	030617-DSDC-E-D-M	10	0.0234	-245.1	PPB	PPB	13 Mar 2017 15:51:16	U	SMPL
I2231-01LX5 - 1		12	0.0240	-245.1	PPB	PPB	13 Mar 2017 15:53:23	U	SMPL
I2231-01A - 1		13311	3.8745	-245.1	PPB	PPB	13 Mar 2017 15:55:31	U	SMPL
I2224-01 - 1	030917-PRRT-F-U-M	25	0.0278	-245.1	PPB	PPB	13 Mar 2017 15:57:39	U	SMPL
I2190-01 - 1	030617-DSDC-E-D-M	76	0.0425	-245.1	PPB	PPB	13 Mar 2017 15:59:47	U	SMPL
CCV95 - 1	CCV95	16931	4.9226	-245.1	PPB	PPB	13 Mar 2017 16:01:55	U	SMPL
CCB95 - 1	CCB95	-24	0.0136	-245.1	PPB	PPB	13 Mar 2017 16:04:04	U	SMPL