



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86345

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/13/2017	16:25
2	CAL2	1	Water			7.01	03/13/2017	16:26
3	CAL3	1	Water			10.03	03/13/2017	16:27
4	ICV	1	Water			7.01	03/13/2017	16:28
5	CCV1	1	Water			2.02	03/13/2017	16:30
6	I2258-01	1	Solid	20.03	20	5.98	03/13/2017	16:31
7	I2258-01DUP	1	Solid	20.05	20	5.99	03/13/2017	16:32
8	CCV2	1	Water			12.01	03/13/2017	16:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2258

WorkList ID : 96494

Date : 3/13/2017 4:10:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2258-01	pH	Cool 4 deg C	TULL02	L31	GRINNELL-TOPSOIL	03/13/2017	9045C

Date/Time 03-13-17 4:10 PM
 Received by: JH
 Relinquished by: CH

Date/Time 03-13-17 5:40 PM
 Received by: CP
 Relinquished by: SP