



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86361

BalanceID: NA

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP53487
COD CCV std, 50ppm	WP53488

Temp In (C): <u>150</u>	Date In: <u>03/15/2017</u>	Time In: <u>09:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>03/15/2017</u>	Time Out: <u>11:00</u>

Intercept: -0.7821

Slope: 1.0223

Regression: 0.9998

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	02/20/2017	14:05
2	CAL2	10	1	Water	9.000	02/20/2017	14:05
3	CAL3	50	1	Water	49.000	02/20/2017	14:06
4	CAL4	100	1	Water	103.000	02/20/2017	14:06
5	CAL5	150	1	Water	152.000	02/20/2017	14:07

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: ASTM COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86361

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	48.696	02/20/2017	14:07
2	ICB				1	Water	0.000	0.765	02/20/2017	14:07
3	CCV1	50			1	Water	53.000	52.609	03/15/2017	11:45
4	CCB1				1	Water	0.000	0.765	03/15/2017	11:45
5	LB86361BL				1	Water	0.000	0.765	03/15/2017	11:46
6	LB86361BS	50			1	Water	51.000	50.653	03/15/2017	11:46
7	I2173-01				1	Water	27.000	27.176	03/15/2017	11:47
8	I2173-01DUP				1	Water	29.000	29.132	03/15/2017	11:47
9	I2173-01MS	50			1	Water	78.000	77.064	03/15/2017	11:48
10	I2173-01MSD	50			1	Water	79.000	78.042	03/15/2017	11:48
11	I2173-02				1	Water	19.000	19.351	03/15/2017	11:49
12	I2173-03				1	Water	26.000	26.198	03/15/2017	11:49
13	I2173-04				1	Water	20.000	20.329	03/15/2017	12:00
14	I2173-05				1	Water	13.000	13.481	03/15/2017	12:00
15	CCV2	50			1	Water	53.000	52.609	03/15/2017	12:01
16	CCB2				1	Water	0.000	0.765	03/15/2017	12:01
17	I2173-06				1	Water	36.000	35.980	03/15/2017	12:02
18	I2173-07				1	Water	26.000	26.198	03/15/2017	12:02
19	I2173-08				1	Water	33.000	33.045	03/15/2017	12:03
20	I2173-09				1	Water	27.000	27.176	03/15/2017	12:03
21	I2173-10				1	Water	26.000	26.198	03/15/2017	12:04
22	I2175-01				1	Water	20.000	20.329	03/15/2017	12:04
23	I2175-01DUP				1	Water	21.000	21.307	03/15/2017	12:05
24	I2175-02				1	Water	19.000	19.351	03/15/2017	12:05
25	I2175-03				1	Water	51.000	50.653	03/15/2017	12:06
26	I2175-04				1	Water	23.000	23.263	03/15/2017	12:06
27	CCV3	50			1	Water	52.000	51.631	03/15/2017	12:07
28	CCB3				1	Water	0.000	0.765	03/15/2017	12:07
29	I2175-05				1	Water	45.000	44.783	03/15/2017	12:08
30	I2175-06				1	Water	62.000	61.413	03/15/2017	12:08
31	I2175-07				1	Water	14.000	14.460	03/15/2017	12:09
32	I2175-08				1	Water	19.000	19.351	03/15/2017	12:09
33	I2175-09				1	Water	47.000	46.740	03/15/2017	12:10
34	CCV4	50			1	Water	52.000	51.631	03/15/2017	12:10
35	CCB4				1	Water	0.000	0.765	03/15/2017	12:11

WORKLIST(Hardcopy Internal Chain)

2886361

WorkList Name : COD031317 WorkList ID : 96493

Date : 3/13/2017 3:38:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2173-01	ASTM COD	Cool 4 deg C	SEVE02	K61	D-38	03/07/2017	SM5220 D
	Solid	I2173-02	ASTM COD	Cool 4 deg C	SEVE02	K61	D-41	03/07/2017	SM5220 D
	Solid	I2173-03	ASTM COD	Cool 4 deg C	SEVE02	K61	D-44	03/07/2017	SM5220 D
	Solid	I2173-04	ASTM COD	Cool 4 deg C	SEVE02	K61	D-47	03/07/2017	SM5220 D
	Solid	I2173-05	ASTM COD	Cool 4 deg C	SEVE02	K61	D-49	03/07/2017	SM5220 D
	Solid	I2173-06	ASTM COD	Cool 4 deg C	SEVE02	K61	D-51	03/07/2017	SM5220 D
	Solid	I2175-09	ASTM COD	Cool 4 deg C	SEVE02	K61	D-85	03/07/2017	SM5220 D
	Solid	I2175-03	ASTM COD	Cool 4 deg C	SEVE02	K61	D-69	03/07/2017	SM5220 D
	Solid	I2175-04	ASTM COD	Cool 4 deg C	SEVE02	K61	D-71	03/07/2017	SM5220 D
	Solid	I2175-05	ASTM COD	Cool 4 deg C	SEVE02	K61	D-74	03/07/2017	SM5220 D
	Solid	I2175-06	ASTM COD	Cool 4 deg C	SEVE02	K61	D-77	03/07/2017	SM5220 D
	Solid	I2175-07	ASTM COD	Cool 4 deg C	SEVE02	K61	D-79	03/07/2017	SM5220 D
	Solid	I2175-08	ASTM COD	Cool 4 deg C	SEVE02	K61	D-81	03/07/2017	SM5220 D
	Solid	I2173-07	ASTM COD	Cool 4 deg C	SEVE02	K61	D-53	03/07/2017	SM5220 D
	Solid	I2173-08	ASTM COD	Cool 4 deg C	SEVE02	K61	D-55	03/07/2017	SM5220 D
	Solid	I2173-09	ASTM COD	Cool 4 deg C	SEVE02	K61	D-57	03/07/2017	SM5220 D
	Solid	I2173-10	ASTM COD	Cool 4 deg C	SEVE02	K61	D-61	03/07/2017	SM5220 D
	Solid	I2175-01	ASTM COD	Cool 4 deg C	SEVE02	K61	D-64	03/07/2017	SM5220 D
	Solid	I2175-02	ASTM COD	Cool 4 deg C	SEVE02	K61	D-67	03/07/2017	SM5220 D

Date/Time 03/15/17 8:40
 Received by: IB
 Relinquished by: CB

Date/Time 03/15/17 11:00
 Received by: IB
 Relinquished by: IB