

7471A


~~LB 86360~~ LB 86365
INSTRUMENT ID: CV1

m3
3/16/17

Linear

 μ Abs.:

46097



A= 0.0000e+000

B= 2.1735e-004

C= 2.0729e-003

Rho= 0.9998583

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	51	0.000	51				
0.2	0.200	0.209	0.009	950	0.0 %	950				
2.5	2.500	2.430	-0.070	11170	0.0 %	11170				
5.0	5.000	5.106	0.106	23484	0.0 %	23484				
7.5	7.500	7.421	-0.079	34134	0.0 %	34134				
10.0	10.000	10.021	0.021	46097	0.0 %	46097				

LB86365 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Mar 2017 12:06:17

Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7471A	PPB	15 Mar 2017 12:08:02	S	Std
0.2 - 1	50.2	950	-	0.2000	7471A	PPB	15 Mar 2017 12:10:20	S	Std
2.5 - 1	50.5	11170	-	2.5000	7471A	PPB	15 Mar 2017 12:12:37	S	Std
5.0 - 1	55	23484	-	5.0000	7471A	PPB	15 Mar 2017 12:14:55	S	Std
7.5 - 1	57.5	34134	-	7.5000	7471A	PPB	15 Mar 2017 12:17:12	S	Std
10.0 - 1	510	46097	-	10.0000	7471A	PPB	15 Mar 2017 12:19:29	S	Std
ICV71 - 1	ICV71	17893	3.8911	-	7471A	PPB	15 Mar 2017 12:22:16	U	SMPL
ICB71 - 1	ICB71	-13	-0.0008	-	7471A	PPB	15 Mar 2017 12:24:31	U	SMPL
CCV04 - 1	CCV04	23150	5.0337	-	7471A	PPB	15 Mar 2017 12:26:47	U	SMPL
CCB04 - 1	CCB04	-30	-0.0044	-	7471A	PPB	15 Mar 2017 12:29:03	U	SMPL
CRA - 1	CRA	916	0.2012	-	7471A	PPB	15 Mar 2017 12:31:19	U	SMPL
HighStd - 1	HighStd	44908	9.7627	-	7471A	PPB	15 Mar 2017 12:33:34	U	SMPL
ChkStd - 1	ChkStd	32388	7.0415	-	7471A	PPB	15 Mar 2017 12:35:51	U	SMPL
PB97584BL - 1	PBS	-45	-0.0077	-	7471A	PPB	15 Mar 2017 12:38:07	U	SMPL
PB97584BS - 1	LCSS	15872	3.4518	-	7471A	PPB	15 Mar 2017 12:40:23	U	SMPL
I2221-01 - 1	HY-SB424	18620	4.0491	-	7471A	PPB	15 Mar 2017 12:42:40	U	SMPL
I2233-01 - 1	FRAC-TANK-BOTTOM	36676	7.9735	-	7471A	PPB	15 Mar 2017 12:44:57	U	SMPL
I2244-02 - 1	TEC-COMP	23590	5.1293	-	7471A	PPB	15 Mar 2017 12:47:15	U	SMPL
I2258-01 - 1	GRINNELL-TOPSOIL	6079	1.3233	-	7471A	PPB	15 Mar 2017 12:49:33	U	SMPL
I2258-01DUP - 1	GRINNELL-TOPSOILDUP	6348	1.3818	-	7471A	PPB	15 Mar 2017 12:51:49	U	SMPL
CCV05 - 1	CCV05	23172	5.0385	-	7471A	PPB	15 Mar 2017 12:54:05	U	SMPL
CCB05 - 1	CCB05	5	0.0032	-	7471A	PPB	15 Mar 2017 12:56:21	U	SMPL
I2258-01MS - 1	GRINNELL-TOPSOILMS	21383	4.6496	-	7471A	PPB	15 Mar 2017 12:58:36	U	SMPL
I2258-01MSD - 1	GRINNELL-TOPSOILMSD	21836	4.7481	-	7471A	PPB	15 Mar 2017 13:00:52	U	SMPL
I2261-01 - 1	TRAP-ROCK-FINES	80	0.0195	-	7471A	PPB	15 Mar 2017 13:03:08	U	SMPL
PB97587BL - 1	PBS	8	0.0038	-	7471A	PPB	15 Mar 2017 13:05:25	U	SMPL
PB97587BS - 1	LCSS	17989	3.9119	-	7471A	PPB	15 Mar 2017 13:07:42	U	SMPL
I2264-01 - 1	E2-BB7-WSW	738	0.1625	-	7471A	PPB	15 Mar 2017 13:09:59	U	SMPL
I2264-02 - 1	E2-BB7-BASE	582	0.1286	-	7471A	PPB	15 Mar 2017 13:12:16	U	SMPL
I2264-03 - 1	E2-BB7-SSW	147014	31.9552	-	7471A	PPB	15 Mar 2017 13:14:34	U	SMPL
I2264-04 - 1	E2-BB6-SSW	227585	49.4672	-	7471A	PPB	15 Mar 2017 13:16:51	U	SMPL
I2264-05 - 1	E2-CC7-BASE	22184	4.8237	-	7471A	PPB	15 Mar 2017 13:19:09	U	SMPL
CCV06 - 1	CCV06	23497	5.1091	-	7471A	PPB	15 Mar 2017 13:21:28	U	SMPL
CCB06 - 1	CCB06	-7	0.0006	-	7471A	PPB	15 Mar 2017 13:23:44	U	SMPL
I2264-06 - 1	E2-CC6-BASE	24298	5.2832	-	7471A	PPB	15 Mar 2017 13:26:00	U	SMPL
I2264-07 - 1	E2-CC6-WSW	51802	11.2611	-	7471A	PPB	15 Mar 2017 13:28:17	U	SMPL
I2264-07DUP - 1	E2-CC6-WSWDUP	47383	10.3007	-	7471A	PPB	15 Mar 2017 13:30:33	U	SMPL
I2264-08 - 1	E2-DD7-BASE	6104	1.3288	-	7471A	PPB	15 Mar 2017 13:32:50	U	SMPL
I2264-09 - 1	E2-DD8-BASE	22970	4.9946	-	7471A	PPB	15 Mar 2017 13:35:07	U	SMPL
I2264-10 - 1	E2-DD7-SSW	412757	HIGH	-	7471A	PPB	15 Mar 2017 13:37:23	X	
I2264-11 - 1	E2-DD8-SSW	114499	24.8882	-	7471A	PPB	15 Mar 2017 13:39:41	U	SMPL
I2264-12 - 1	E2-DD6-WSW	204288	44.4036	-	7471A	PPB	15 Mar 2017 13:41:58	U	SMPL
CCV07 - 1	CCV07	23038	5.0093	-	7471A	PPB	15 Mar 2017 13:43:51	U	SMPL
CCB07 - 1	CCB07	-13	-0.0008	-	7471A	PPB	15 Mar 2017 13:45:09	U	SMPL
I2264-13 - 1	E2-DD9-BASE	377439	HIGH	-	7471A	PPB	15 Mar 2017 13:47:27	X	
I2264-14 - 1	E2-EE8-WSW	296099	HIGH	-	7471A	PPB	15 Mar 2017 13:49:44	X	
I2264-15 - 1	E2-EE8-SSW	132178	28.7307	-	7471A	PPB	15 Mar 2017 13:51:02	U	SMPL
I2264-16 - 1	E2-EE9-ESW	34063	7.4056	-	7471A	PPB	15 Mar 2017 13:53:19	U	SMPL
I2264-17 - 1	E2-EE9-BASE	63470	13.7971	-	7471A	PPB	15 Mar 2017 13:55:36	U	SMPL
I2264-18 - 1	E2-X6-BASE	45451	9.8807	-	7471A	PPB	15 Mar 2017 14:01:22	U	SMPL
I2266-01 - 1	TP6	4866	1.0597	-	7471A	PPB	15 Mar 2017 14:03:39	U	SMPL
I2266-02 - 1	TP7	24159	5.2530	-	7471A	PPB	15 Mar 2017 14:05:56	U	SMPL
I2264-07MS - 1	E2-CC6-WSWMS	70506	15.3264	-	7471A	PPB	15 Mar 2017 14:08:13	U	SMPL
I2264-07MSD - 1	E2-CC6-WSWMSD	66846	14.5309	-	7471A	PPB	15 Mar 2017 14:10:30	U	SMPL
CCV08 - 1	CCV08	23281	5.0621	-	7471A	PPB	15 Mar 2017 14:12:47	U	SMPL
CCB08 - 1	CCB08	-11	-0.0003	-	7471A	PPB	15 Mar 2017 14:15:04	U	SMPL
I2258-01LX5 - 1		1312	0.2872	-	7471A	PPB	15 Mar 2017 14:17:21	U	SMPL
I2258-01A - 1		23743	5.1626	-	7471A	PPB	15 Mar 2017 14:19:38	U	SMPL
I2264-07LX5 - 1		10280	2.2364	-	7471A	PPB	15 Mar 2017 14:21:55	U	SMPL
I2264-07A - 1		68032	14.7887	-	7471A	PPB	15 Mar 2017 14:24:12	U	SMPL
I2264-10DLX10 - 1		57797	12.5641	-	7471A	PPB	15 Mar 2017 14:26:29	U	SMPL
I2264-13DLX10 - 1		53846	11.7054	-	7471A	PPB	15 Mar 2017 14:28:46	U	SMPL
I2264-14DLX10 - 1		39914	8.6773	-	7471A	PPB	15 Mar 2017 14:31:03	U	SMPL
I2264-03DLX5 - 1		30765	6.6888	-	7471A	PPB	15 Mar 2017 14:33:20	U	SMPL
I2264-04DLX10 - 1		24556	5.3393	-	7471A	PPB	15 Mar 2017 14:35:37	U	SMPL
I2264-11DLX5 - 1		23207	5.0461	-	7471A	PPB	15 Mar 2017 14:37:54	U	SMPL
CCV09 - 1	CCV09	23656	5.1437	-	7471A	PPB	15 Mar 2017 14:40:11	U	SMPL
CCB09 - 1	CCB09	5	0.0032	-	7471A	PPB	15 Mar 2017 14:42:28	U	SMPL
I2264-12DLX10 - 1		21343	4.6409	-	7471A	PPB	15 Mar 2017 14:44:45	U	SMPL
I2264-15DLX5 - 1		27397	5.9567	-	7471A	PPB	15 Mar 2017 14:47:02	U	SMPL
I2264-17DLX5 - 1		12744	2.7720	-	7471A	PPB	15 Mar 2017 14:49:19	U	SMPL
I2264-07DLX5 - 1		10272	2.2347	-	7471A	PPB	15 Mar 2017 14:51:36	U	SMPL
I2264-07DLX5DUP - 1		9523	2.0719	-	7471A	PPB	15 Mar 2017 14:53:53	U	SMPL
I2264-07DLX5MS - 1		14151	3.0778	-	7471A	PPB	15 Mar 2017 14:56:10	U	SMPL
I2264-07DLX5MSD - 1		13443	2.9239	-	7471A	PPB	15 Mar 2017 14:58:27	U	SMPL
I2264-07LDLX25 - 1		2091	0.4565	-	7471A	PPB	15 Mar 2017 15:00:44	U	SMPL
I2264-07ADLX5 - 1		13784	2.9980	-	7471A	PPB	15 Mar 2017 15:02:61	U	SMPL
CCV10 - 1	CCV10	23219	5.0487	-	7471A	PPB	15 Mar 2017 15:04:78	U	SMPL
CCB10 - 1	CCB10	6	0.0034	-	7471A	PPB	15 Mar 2017 15:06:95	U	SMPL
I2264-10DL2X5 - 1		11437	2.4879	-	7471A	PPB	15 Mar 2017 15:09:12	U	SMPL
I2264-13DL2X5 - 1		10366	2.2551	-	7471A	PPB	15 Mar 2017 15:11:29	U	SMPL

LB86365 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Mar 2017 12:06:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCV11 - 1	CCV11	23330	5.0728		7471A	PPB	15 Mar 2017 15:35:14	U	SMPL
CCB11 - 1	CCB11	8	0.0038		7471A	PPB	15 Mar 2017 15:37:30	U	SMPL