



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 03/16/2017Run Number: LB86373DATE & TIME IN: 03/15/2017 16:30OVEN TEMP: 98 °CBalanceID: WC SC-3DATE & TIME OUT: 03/16/2017 12:45OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @105°C (g)	2nd Dish+Sample weight after Drying @105°C (g)	Final Dish+Sample weight after Drying @105°C (g)	Weight (g)	Result (mg/L)
1	LB86373BL	LB86373BL	95.8026	95.8026	100	95.8026	95.8026	95.8026	0	0
2	I2242-01	TEC-GW	90.8353	90.8353	100	90.9861	90.9861	90.9861	0.1508	1508
3	I2242-01DUP	TEC-GWDUP	91.7403	91.7403	100	91.8914	91.8914	91.8914	0.1511	1511

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2B86343

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS 12242 WorkList ID : 96546 Date : 3/15/2017 3:59:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2242-01	TS	Cool 4 deg C	TECT01	L31	TEC-GW	03/10/2017	SM2540 B

Date/Time 03.15.17 3:59:01 PM
Received by: JP
Relinquished by: CA

Date/Time 03.15.17 5:35 PM
Received by: AP
Relinquished by: JP