


7470A

LB 86379
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36786



A= 0.0000e+000

B= 2.7050e-004

C= 1.1402e-002

Rho= 0.9998436

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	14	0.000	14				
0.2	0.200	0.194	-0.006	676	0.0 %	676				
2.5	2.500	2.525	0.025	9291	0.0 %	9291				
5.0	5.000	4.894	-0.106	18051	0.0 %	18051				
7.5	7.500	7.610	0.110	28090	0.0 %	28090				
10.0	10.000	9.962	-0.038	36786	0.0 %	36786				

LB86377 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Mar 2017 10:40:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	14	-	0.0000	7470A	PPB	16 Mar 2017 10:48:46	S	Std
0.2 - 1	S0-2	676	-	0.2000	7470A	PPB	16 Mar 2017 10:50:55	S	Std
2.5 - 1	S2.5	9291	-	2.5000	7470A	PPB	16 Mar 2017 10:53:03	S	Std
5.0 - 1	S5	18051	-	5.0000	7470A	PPB	16 Mar 2017 10:55:11	S	Std
7.5 - 1	S7.5	28090	-	7.5000	7470A	PPB	16 Mar 2017 10:57:19	S	Std
10.0 - 1	S10	36786	-	10.0000	7470A	PPB	16 Mar 2017 10:59:26	S	Std
ICV72 - 1	ICV72	15261	4.1395	-7470A	PPB	16 Mar 2017 11:04:29	U	SMPL	
ICB72 - 1	ICB72	-72	-0.0081	-7470A	PPB	16 Mar 2017 11:06:35	U	SMPL	
CCV12 - 1	CCV12	17654	4.7868	-7470A	PPB	16 Mar 2017 11:08:41	U	SMPL	
CCB12 - 1	CCB12	-95	-0.0143	-7470A	PPB	16 Mar 2017 11:10:48	U	SMPL	
CRA - 1	CRA	653	0.1880	-7470A	PPB	16 Mar 2017 11:12:54	U	SMPL	
HighStd - 1	HighStd	36596	9.9106	-7470A	PPB	16 Mar 2017 11:15:01	U	SMPL	
ChkStd - 1	ChkStd	25600	6.9362	-7470A	PPB	16 Mar 2017 11:17:08	U	SMPL	
PB97618BL - 1	PBW	-120	-0.0211	-7470A	PPB	16 Mar 2017 11:19:16	U	SMPL	
PB97618BS - 1	LCSW	14339	3.8901	-7470A	PPB	16 Mar 2017 11:21:24	U	SMPL	
PB97618TB - 1	PB97582TB	-71	-0.0078	-7470A	PPB	16 Mar 2017 11:23:31	U	SMPL	
I2244-03 - 1	TEC-COMP	11	0.0144	-7470A	PPB	16 Mar 2017 11:25:40	U	SMPL	
I2244-03DUP - 1	TEC-COMPDUP	11	0.0144	-7470A	PPB	16 Mar 2017 11:27:48	U	SMPL	
I2244-03MS - 1	TEC-COMPMS	13281	3.6039	-7470A	PPB	16 Mar 2017 11:29:57	U	SMPL	
I2244-03MSD - 1	TEC-COMPMSD	13413	3.6396	-7470A	PPB	16 Mar 2017 11:32:04	U	SMPL	
CCV13 - 1	CCV13	17994	4.8788	-7470A	PPB	16 Mar 2017 11:34:11	U	SMPL	
CCB13 - 1	CCB13	-82	-0.0108	-7470A	PPB	16 Mar 2017 11:36:17	U	SMPL	
PB97619BL - 1	PBW	-1	0.0111	-7470A	PPB	16 Mar 2017 11:38:24	U	SMPL	
PB97619BS - 1	LCSW	14520	3.9391	-7470A	PPB	16 Mar 2017 11:40:30	U	SMPL	
I2210-01 - 1	FB-2017-03-07	-7	0.0095	-7470A	PPB	16 Mar 2017 11:42:37	U	SMPL	
I2264-19 - 1	E2-FB-3-10-17	105	0.0398	-7470A	PPB	16 Mar 2017 11:44:44	U	SMPL	
I2244-03LX5 - 1		19	0.0165	-7470A	PPB	16 Mar 2017 11:49:30	U	SMPL	
I2244-03A - 1		15422	4.1830	-7470A	PPB	16 Mar 2017 11:51:37	U	SMPL	
CCV14 - 1	CCV14	18087	4.9039	-7470A	PPB	16 Mar 2017 11:53:44	U	SMPL	
CCB14 - 1	CCB14	-76	-0.0092	-7470A	PPB	16 Mar 2017 11:55:53	U	SMPL	