



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86378

Temp.: 22.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/20/2017	09:00
2	CAL2	1	Water			7.00	03/20/2017	09:01
3	CAL3	1	Water			10.02	03/20/2017	09:02
4	ICV	1	Water			7.00	03/20/2017	09:03
5	CCV1	1	Water			2.01	03/20/2017	09:04
6	I2212-20	1	Water			9.54	03/20/2017	09:05
7	CCV2	1	Water			12.02	03/20/2017	09:06