


245.1

LB86381
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33904



A= 0.0000e+000

B= 2.9622e-004

C= 2.8449e-002

Rho= 0.9999112

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	9	0.000	9				
0.2	0.200	0.216	0.016	632	0.0 %	632				
2.5	2.500	2.471	-0.029	8246	0.0 %	8246				
5.0	5.000	4.996	-0.004	16770	0.0 %	16770				
7.5	7.500	7.414	-0.086	24934	0.0 %	24934				
10.0	10.000	10.072	0.072	33904	0.0 %	33904				

LB86381 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 16 Mar 2017 13:07:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	245.1	PPB	16 Mar 2017 14:34:22	S	Std
0.2 - 1	50.2	632	-	0.2000	245.1	PPB	16 Mar 2017 14:36:30	S	Std
2.5 - 1	52.5	8246	-	2.5000	245.1	PPB	16 Mar 2017 14:38:38	S	Std
5.0 - 1	55	16770	-	5.0000	245.1	PPB	16 Mar 2017 14:40:47	S	Std
7.5 - 1	57.5	24934	-	7.5000	245.1	PPB	16 Mar 2017 14:43:29	S	Std
10.0 - 1	510	33904	-	10.0000	245.1	PPB	16 Mar 2017 14:45:36	S	Std
ICV73 - 1	ICV73	13343	3.9809	-245.1		PPB	16 Mar 2017 14:48:13	U	SMPL
ICB73 - 1	ICB73	-55	0.0122	-245.1		PPB	16 Mar 2017 14:50:19	U	SMPL
CCV15 - 1	CCV15	16731	4.9845	-245.1		PPB	16 Mar 2017 14:52:25	U	SMPL
CCB15 - 1	CCB15	-57	0.0116	-245.1		PPB	16 Mar 2017 14:54:32	U	SMPL
CRA - 1	CRA	610	0.2091	-245.1		PPB	16 Mar 2017 14:56:38	U	SMPL
HighStd - 1	HighStd	33949	10.0849	-245.1		PPB	16 Mar 2017 14:58:45	U	SMPL
ChkStd - 1	ChkStd	23092	6.8688	-245.1		PPB	16 Mar 2017 15:00:52	U	SMPL
PB97620BL - 1	PBW	10	0.0314	-245.1		PPB	16 Mar 2017 15:03:00	U	SMPL
PB97620BS - 1	LCSW	14359	4.2819	-245.1		PPB	16 Mar 2017 15:05:07	U	SMPL
I2097-01 - 1	EFF-DAY-1-METAL	385	0.1425	-245.1		PPB	16 Mar 2017 15:11:46	U	SMPL
I2097-01DUP - 1	EFF-DAY-1-METALDUP	330	0.1262	-245.1		PPB	16 Mar 2017 15:13:53	U	SMPL
I2097-01MS - 1	EFF-DAY-1-METALMS	13971	4.1670	-245.1		PPB	16 Mar 2017 15:16:01	U	SMPL
I2097-01MSD - 1	EFF-DAY-1-METALMSD	12775	3.8127	-245.1		PPB	16 Mar 2017 15:18:10	U	SMPL
I2162-01 - 1	031317-PRRT-F-D-M	-4	0.0273	-245.1		PPB	16 Mar 2017 15:20:17	U	SMPL
CCV16 - 1	CCV16	15652	4.6649	-245.1		PPB	16 Mar 2017 15:22:24	U	SMPL
CCB16 - 1	CCB16	-31	0.0193	-245.1		PPB	16 Mar 2017 15:24:31	U	SMPL
I2162-02 - 1	031317-PRRT-F-U-M	40	0.0403	-245.1		PPB	16 Mar 2017 15:26:38	U	SMPL
I2167-01 - 1	031117-DSDC-F-D-M	58	0.0456	-245.1		PPB	16 Mar 2017 15:28:45	U	SMPL
I2167-02 - 1	031117-DSDC-F-U-M	59	0.0459	-245.1		PPB	16 Mar 2017 15:30:52	U	SMPL
I2242-01 - 1	TEC-GW	44	0.0415	-245.1		PPB	16 Mar 2017 15:32:59	U	SMPL
I2097-01LX5 - 1		220	0.0936	-245.1		PPB	16 Mar 2017 15:35:06	U	SMPL
I2097-01A - 1		13207	3.9407	-245.1		PPB	16 Mar 2017 15:37:14	U	SMPL
CCV17 - 1	CCV17	16096	4.7964	-245.1		PPB	16 Mar 2017 15:39:22	U	SMPL
CCB17 - 1	CCB17	-46	0.0148	-245.1		PPB	16 Mar 2017 15:41:30	U	SMPL