

Analysis Method: 9034

ANALYST: ketankumar

Parameter: Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB86392

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2169
IODINE SOLUTION .025N 1L	W2178
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB97633BL		1	50	50	5.00	0.00	4.76	4.76	0.24	0.00	0.00	03/16/2017	13:45
2	PB97633BS	25	1	50	50	5.00	0.00	2.22	2.22	2.78	2.54	20.32	03/16/2017	13:50
3	I2212-32		1	50	50	5.00	0.00	3.88	3.88	1.12	0.88	7.04	03/16/2017	13:55

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume