

7471A


LB86404

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45218



A= 0.0000e+000

B= 2.2311e-004

C= -1.1024e-002

Rho= 0.9998375

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	34	0.000	34				
0.2	0.200	0.211	0.011	994	0.0 %	994				
2.5	2.500	2.515	0.015	11321	0.0 %	11321				
5.0	5.000	5.037	0.037	22628	0.0 %	22628				
7.5	7.500	7.363	-0.137	33051	0.0 %	33051				
10.0	10.000	10.077	0.077	45218	0.0 %	45218				

LB86404 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 17 Mar 2017 10:58:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	34	-	0.0000	7471A	PPB	17 Mar 2017 11:14:45	S	Std
0.2 - 1	SO-2	994	-	0.2000	7471A	PPB	17 Mar 2017 11:17:02	S	Std
2.5 - 1	SO-5	11321	-	2.5000	7471A	PPB	17 Mar 2017 11:19:20	S	Std
5.0 - 1	SO	22628	-	5.0000	7471A	PPB	17 Mar 2017 11:21:37	S	Std
7.5 - 1	SO-5	33051	-	7.5000	7471A	PPB	17 Mar 2017 11:23:55	S	Std
10.0 - 1	SO	45218	-	10.0000	7471A	PPB	17 Mar 2017 11:31:58	S	Std
ICV74 - 1	ICV74	18141	4.0364	-	7471A	PPB	17 Mar 2017 11:34:40	U	SMPL
ICB74 - 1	ICB74	-6	-0.0124	-	7471A	PPB	17 Mar 2017 11:36:55	U	SMPL
CCV18 - 1	CCV18	22552	5.0205	-	7471A	PPB	17 Mar 2017 11:39:11	U	SMPL
CCB18 - 1	CCB18	-17	-0.0148	-	7471A	PPB	17 Mar 2017 11:41:26	U	SMPL
CRA - 1	CRA	989	0.2096	-	7471A	PPB	17 Mar 2017 11:43:42	U	SMPL
HighStd - 1	HighStd	45728	10.1913	-	7471A	PPB	17 Mar 2017 11:45:58	U	SMPL
ChkStd - 1	ChkStd	30255	6.7391	-	7471A	PPB	17 Mar 2017 11:48:14	U	SMPL
PB97645BL - 1	PBS	-20	-0.0155	-	7471A	PPB	17 Mar 2017 11:50:31	U	SMPL
PB97645BS - 1	LCSS	15968	3.5516	-	7471A	PPB	17 Mar 2017 11:52:47	U	SMPL
I2270-02 - 1	41B-9-11	307	0.0575	-	7471A	PPB	17 Mar 2017 12:00:23	U	SMPL
I2270-02DUP - 1	41B-9-11DUP	275	0.0503	-	7471A	PPB	17 Mar 2017 12:02:39	U	SMPL
I2270-02MS - 1	41B-9-11MS	16326	3.6314	-	7471A	PPB	17 Mar 2017 12:04:57	U	SMPL
I2270-02MSD - 1	41B-9-11MSD	16676	3.7095	-	7471A	PPB	17 Mar 2017 12:07:15	U	SMPL
I2271-01 - 1	BHP-1	1386	0.2982	-	7471A	PPB	17 Mar 2017 12:09:31	U	SMPL
CCV19 - 1	CCV19	22551	5.0203	-	7471A	PPB	17 Mar 2017 12:11:47	U	SMPL
CCB19 - 1	CCB19	4	-0.0101	-	7471A	PPB	17 Mar 2017 12:14:03	U	SMPL
I2271-03 - 1	BHP-2	1125	0.2400	-	7471A	PPB	17 Mar 2017 12:16:18	U	SMPL
I2277-01 - 1	S39-9003	252	0.0452	-	7471A	PPB	17 Mar 2017 12:18:34	U	SMPL
I2270-02LX5 - 1		75	0.0057	-	7471A	PPB	17 Mar 2017 12:20:50	U	SMPL
I2270-02A - 1		18276	4.0665	-	7471A	PPB	17 Mar 2017 12:23:07	U	SMPL
CCV20 - 1	CCV20	22705	5.0547	-	7471A	PPB	17 Mar 2017 12:25:24	U	SMPL
CCB20 - 1	CCB20	-2	-0.0115	-	7471A	PPB	17 Mar 2017 12:27:40	U	SMPL