



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86405

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/10/2017	17:30
2	CAL2	1	Water			7.02	03/10/2017	17:31
3	CAL3	1	Water			10.03	03/10/2017	17:32
4	ICV	1	Water			7.01	03/10/2017	17:33
5	CCV1	1	Water			2.02	03/10/2017	17:34
6	I2244-03	1	Solid	20.03	20	9.01	03/10/2017	17:35
7	I2244-03DUP	1	Solid	20.05	20	9.02	03/10/2017	17:36
8	CCV2	1	Water			12.02	03/10/2017	17:37

20386405

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2244 WorkList ID : 96610 Date : 3/10/2017 5:10:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2244-03	Corrosivity	Cool 4 deg C	TECT01	L31	TEC-COMP	03/10/2017	9045C

Date/Time 03-10-17 5:10 PM
Received by: JQ
Relinquished by: CQ

Date/Time 03-10-17 5:00 PM
Received by: CP
Relinquished by: JQ