

7470A


LB 86407

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36753



A= 0.0000e+000

B= 2.7184e-004

C= -1.7298e-002

Rho= 0.9999832

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	34	0.000	34				
0.2	0.200	0.185	-0.015	746	0.0 %	746				
2.5	2.500	2.500	-0.000	9260	0.0 %	9260				
5.0	5.000	5.041	0.041	18609	0.0 %	18609				
7.5	7.500	7.508	0.008	27681	0.0 %	27681				
10.0	10.000	9.974	-0.026	36753	0.0 %	36753				

LB86407 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Mar 2017 12:36:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7470A	PPB	17 Mar 2017 12:39:47	S	Std
0.2 - 1	50-2	746	-	0.2000	7470A	PPB	17 Mar 2017 12:41:55	S	Std
2.5 - 1	52.5	9260	-	2.5000	7470A	PPB	17 Mar 2017 12:44:03	S	Std
5.0 - 1	55	18609	-	5.0000	7470A	PPB	17 Mar 2017 12:46:12	S	Std
7.5 - 1	57.5	27681	-	7.5000	7470A	PPB	17 Mar 2017 12:48:20	S	Std
10.0 - 1	510	36753	-	10.0000	7470A	PPB	17 Mar 2017 12:52:53	S	Std
ICV75 - 1	ICV75	13991	3.7860	-	7470A	PPB	17 Mar 2017 12:55:26	U	SMPL
ICB75 - 1	ICB75	-46	-0.0298	-	7470A	PPB	17 Mar 2017 12:57:32	U	SMPL
CCV21 - 1	CCV21	17896	4.8476	-	7470A	PPB	17 Mar 2017 12:59:38	U	SMPL
CCB21 - 1	CCB21	-42	-0.0287	-	7470A	PPB	17 Mar 2017 13:01:45	U	SMPL
CRA - 1	CRA	723	0.1792	-	7470A	PPB	17 Mar 2017 13:03:51	U	SMPL
HighStd	HighStd	36968	10.0321	-	7470A	PPB	17 Mar 2017 13:05:58	U	SMPL
ChkStd - 1	ChkStd	26166	7.0957	-	7470A	PPB	17 Mar 2017 13:08:05	U	SMPL
PB97643BL - 1	PBW	-79	-0.0388	-	7470A	PPB	17 Mar 2017 13:10:13	U	SMPL
PB97643BS - 1	LCSW	13533	3.6615	-	7470A	PPB	17 Mar 2017 13:12:20	U	SMPL
I2263-01 - 1	FB-03132017	-8	-0.0195	-	7470A	PPB	17 Mar 2017 13:14:28	U	SMPL
I2263-02 - 1	MW-101S	243	0.0488	-	7470A	PPB	17 Mar 2017 13:16:36	U	SMPL
I2263-03 - 1	MW-101I	525	0.1254	-	7470A	PPB	17 Mar 2017 13:18:44	U	SMPL
I2263-04 - 1	MW-101D	162	0.0267	-	7470A	PPB	17 Mar 2017 13:20:52	U	SMPL
I2263-05 - 1	MW-102S	60	-0.0010	-	7470A	PPB	17 Mar 2017 13:22:59	U	SMPL
CCV22 - 1	CCV22	17624	4.7736	-	7470A	PPB	17 Mar 2017 13:25:06	U	SMPL
CCB22 - 1	CCB22	-34	-0.0265	-	7470A	PPB	17 Mar 2017 13:27:12	U	SMPL
I2263-05DUP - 1	MW-102SDUP	46	-0.0048	-	7470A	PPB	17 Mar 2017 13:29:19	U	SMPL
I2263-06 - 1	I2263-05MS	14166	3.8336	-	7470A	PPB	17 Mar 2017 13:31:26	U	SMPL
I2263-07 - 1	I2263-05MSD	13795	3.7328	-	7470A	PPB	17 Mar 2017 13:33:33	U	SMPL
I2263-08 - 1	MW-102I	-36	-0.0271	-	7470A	PPB	17 Mar 2017 13:35:40	U	SMPL
I2263-09 - 1	MW-102D	129	0.0178	-	7470A	PPB	17 Mar 2017 13:37:47	U	SMPL
I2263-10 - 1	MW-107S	222	0.0431	-	7470A	PPB	17 Mar 2017 13:39:55	U	SMPL
I2263-12 - 1	MW-105S	583	0.1412	-	7470A	PPB	17 Mar 2017 13:42:03	U	SMPL
I2263-13 - 1	MW-105I	100	0.0099	-	7470A	PPB	17 Mar 2017 13:44:11	U	SMPL
I2263-14 - 1	MW-105D	179	0.0314	-	7470A	PPB	17 Mar 2017 13:46:20	U	SMPL
I2263-15 - 1	MW-106S	90	0.0072	-	7470A	PPB	17 Mar 2017 13:48:27	U	SMPL
CCV23 - 1	CCV23	17528	4.7475	-	7470A	PPB	17 Mar 2017 13:50:33	U	SMPL
CCB23 - 1	CCB23	-24	-0.0238	-	7470A	PPB	17 Mar 2017 13:52:40	U	SMPL
I2263-16 - 1	MW-106I	145	0.0221	-	7470A	PPB	17 Mar 2017 13:54:47	U	SMPL
I2263-17 - 1	MW-106D	161	0.0265	-	7470A	PPB	17 Mar 2017 13:56:54	U	SMPL
I2263-05LX5 - 1		37	-0.0072	-	7470A	PPB	17 Mar 2017 13:59:00	U	SMPL
I2263-05A - 1		13159	3.5599	-	7470A	PPB	17 Mar 2017 14:01:07	U	SMPL
CCV24 - 1	CCV24	17645	4.7793	-	7470A	PPB	17 Mar 2017 14:03:15	U	SMPL
CCB24 - 1	CCB24	-40	-0.0282	-	7470A	PPB	17 Mar 2017 14:05:23	U	SMPL