



Analytical Summary Report

Analysis Method:	SM4500-O G	Reviewed By:	Indiana
Parameter:	Dissolved Oxygen	Supervisor Review By:	apatel
Run Number:	LB86408	MANGANOUS SULFATE SOLUTION:	W2191
BOD Water:	WP535228	Alkaline Iodide Azide:	W2035
Starch:	W2121	Sodium Thiosulfate, 0.025N:	W2169
Conc Sulfuric Acid:	W2151	pH Strips:	N/A
		Zero DO Std:	WP53543

Seq	LabID	ClientID	Sample Vol (ml)	Initial Reading (mL)	Final Reading (mL)	Diff (mg/L)	Result (mg/L)
1	WINKLER1	WINKLER1	300	0.00	8.60	8.60	8.50
2	WINKLER2	WINKLER2	300	8.60	17.00	8.40	

Calibrate DO meter with average reading : 8.50

Zero DO Reading : 0.21 Calibrated to : 0.00

Seq	LabID	ClientID	DL	Sample Vol (ml)	Temperature (°C)	Result (mg/L)	Anal Date	Anal Time
1	LB86408BL	LB86408BL	1	300	19.0	0.00	03/16/2017	12:20
2	I2297-01	MW001	1	300	18.5	5.71	03/16/2017	12:23
3	I2297-01DUP	MW001DUP	1	300	18.6	5.70	03/16/2017	12:26
4	I2297-05	SW-02	1	300	18.2	6.38	03/16/2017	12:29

WORKLIST(Hardcopy Internal Chain)

WorkList Name : DO031617 WorkList ID : 96616 Date : 3/16/2017 11:40:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2297-01	Dissolved Oxygen	Cool 4 deg C	PWGR01	L62	MW001	03/15/2017	SM4500-O G
	Water	I2297-05	Dissolved Oxygen	Cool 4 deg C	PWGR01	L62	SW-02	03/15/2017	SM4500-O G

Date/Time 03/16/17 11:50
 Received by: IB
 Relinquished by: CP

Date/Time 03/16/17 13:00
 Received by: CP
 Relinquished by: IB