



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86410

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/17/2017	12:30
2	I2242-01		1	>100.0	>212.0	03/17/2017	13:00
3	I2242-01DUP		1	>100.0	>212.0	03/17/2017	13:32

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86410

WorkList Name : FP31717

WorkList ID : 96612

Date : 3/17/2017 12:05:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12242-01	Flash Point	Cool 4 deg C	TECT01	L31	TEC-GW	03/10/2017	1010A

Date/Time

3/17/17 12:10 PM

Received by:

KS

Relinquished by:

KS

Date/Time

3/17/17 13:45

Received by:

KS

Relinquished by:

KS