



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86415

Temp.: 16.3 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	03/17/2017	13:20
2	CAL2	1	Water			7.01	03/17/2017	13:21
3	CAL3	1	Water			10.02	03/17/2017	13:22
4	ICV	1	Water			6.99	03/17/2017	13:23
5	CCV1	1	Water			2.02	03/17/2017	13:24
6	I2310-01	1	Water			6.15	03/17/2017	13:25
7	I2310-01DUP	1	Water			6.16	03/17/2017	13:26
8	CCV2	1	Water			12.02	03/17/2017	13:27

1086415

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2310 WorkList ID : 96619 Date : 3/17/2017 1:07:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2017	Water	I2310-01	Corrosivity	Cool 4 deg C	HOME01	K51	BASEMENT	03/12/2017	9040C

Date/Time 03-17-17 1:10 PM
 Received by: JP
 Relinquished by: CP

Date/Time 03-17-17 2:30 PM
 Received by: CP
 Relinquished by: JP