


245.1

LB 86428
INSTRUMENT ID: CV4

Linear

 μ Abs.:

33211



A= 0.0000e+000

B= 3.0428e-004

C= -6.2715e-002

Rho= 0.9999636

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	257	0.000	257				
0.2	0.200	0.202	0.002	869	0.0 %	869				
2.5	2.500	2.521	0.021	8492	0.0 %	8492				
5.0	5.000	4.951	-0.049	16478	0.0 %	16478				
7.5	7.500	7.468	-0.032	24748	0.0 %	24748				
10.0	10.000	10.043	0.043	33211	0.0 %	33211				

LB86428 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Mar 2017 16:10:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	257	-	0.0000	245.1	PPB	17 Mar 2017 16:12:43	S	Std
0.2 - 1	50.2	869	-	0.2000	245.1	PPB	17 Mar 2017 16:14:51	S	Std
2.5 - 1	52.5	8492	-	2.5000	245.1	PPB	17 Mar 2017 16:17:00	S	Std
5.0 - 1	55	16478	-	5.0000	245.1	PPB	17 Mar 2017 16:19:08	S	Std
7.5 - 1	57.5	24748	-	7.5000	245.1	PPB	17 Mar 2017 16:21:16	S	Std
10.0 - 1	510	33211	-	10.0000	245.1	PPB	17 Mar 2017 16:23:23	S	Std
ICV76 - 1	ICV76	12841	3.8445	-245.1		PPB	17 Mar 2017 16:26:25	U	SMPL
ICB76 - 1	ICB76	206	-0.0000	-245.1		PPB	17 Mar 2017 16:28:31	U	SMPL
CCV25 - 1	CCV25	16459	4.9454	-245.1		PPB	17 Mar 2017 16:30:38	U	SMPL
CCB25 - 1	CCB25	200	-0.0019	-245.1		PPB	17 Mar 2017 16:32:44	U	SMPL
CRA - 1	CRA	855	0.1974	-245.1		PPB	17 Mar 2017 16:34:51	U	SMPL
HighStd - 1	HighStd	33098	10.0084	-245.1		PPB	17 Mar 2017 16:36:57	U	SMPL
ChkStd - 1	ChkStd	22691	6.8417	-245.1		PPB	17 Mar 2017 16:39:04	U	SMPL
PB97667BL - 1	PBW	-2	-0.0633	-245.1		PPB	17 Mar 2017 16:41:12	U	SMPL
PB97667BS - 1	LCSW	13727	4.1141	-245.1		PPB	17 Mar 2017 16:43:19	U	SMPL
I2196-08 - 1	PT-HG-WP	27390	8.2715	-245.1		PPB	17 Mar 2017 16:47:35	U	SMPL
CCV26 - 1	CCV26	16276	4.8898	-245.1		PPB	17 Mar 2017 16:54:01	U	SMPL
CCB26 - 1	CCB26	186	-0.0061	-245.1		PPB	17 Mar 2017 16:56:08	U	SMPL