

7470A


LB 86432

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36124



A= 0.0000e+000

B= 2.7977e-004

C= -3.3879e-002

Rho= 0.9998798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	210	0.000	210				
0.2	0.200	0.188	-0.012	792	0.0 %	792				
2.5	2.500	2.512	0.012	9099	0.0 %	9099				
5.0	5.000	5.018	0.018	18058	0.0 %	18058				
7.5	7.500	7.385	-0.115	26519	0.0 %	26519				
10.0	10.000	10.072	0.072	36124	0.0 %	36124				

LB86432 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Mar 2017 17:04:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	210	-	0.0000	7470A	PPB	17 Mar 2017 17:08:17	S	Std
0.2 - 1	50.2	792	-	0.2000	7470A	PPB	17 Mar 2017 17:10:25	S	Std
2.5 - 1	52.5	9099	-	2.5000	7470A	PPB	17 Mar 2017 17:12:34	S	Std
5.0 - 1	55	18058	-	5.0000	7470A	PPB	17 Mar 2017 17:14:42	S	Std
7.5 - 1	57.5	26519	-	7.5000	7470A	PPB	17 Mar 2017 17:16:50	S	Std
10.0 - 1	510	36124	-	10.0000	7470A	PPB	17 Mar 2017 17:18:58	S	Std
ICV77 - 1	ICV77	14249	3.9525	-	7470A	PPB	17 Mar 2017 17:21:42	U	SMPL
ICB77 - 1	ICB77	-2	-0.0344	-	7470A	PPB	17 Mar 2017 17:23:48	U	SMPL
CCV27 - 1	CCV27	17909	4.9764	-	7470A	PPB	17 Mar 2017 17:25:54	U	SMPL
CCB27 - 1	CCB27	-3	-0.0347	-	7470A	PPB	17 Mar 2017 17:28:01	U	SMPL
CRA - 1	CRA	761	0.1790	-	7470A	PPB	17 Mar 2017 17:30:07	U	SMPL
HighStd - 1	HighStd	35914	10.0136	-	7470A	PPB	17 Mar 2017 17:32:14	U	SMPL
ChkStd - 1	ChkStd	24865	6.9225	-	7470A	PPB	17 Mar 2017 17:34:21	U	SMPL
PB97668BL - 1	PBW	-22	-0.0400	-	7470A	PPB	17 Mar 2017 17:36:29	U	SMPL
PB97668BS - 1	LCSW	13595	3.7695	-	7470A	PPB	17 Mar 2017 17:38:37	U	SMPL
I2197-08 - 1	PT-HG-WP	28733	8.0046	-	7470A	PPB	17 Mar 2017 17:47:07	U	SMPL
CCV28 - 1	CCV28	18053	5.0167	-	7470A	PPB	17 Mar 2017 17:51:24	U	SMPL
CCB28 - 1	CCB28	111	-0.0028	-	7470A	PPB	17 Mar 2017 17:53:31	U	SMPL