



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB86444

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP53621
Turbidity std, 40NTU	WP53615
turbidity std 5 NTU	WP53618
turbidity std 20 NTU	WP53616
Turbidity std 1 NTU	WP53619
turbidity std 0 NTU	WP53620
turbidity std 10 NTU	WP53617
10 NTU Standard 500 ml	W2152

Intercept: 0.0006

Slope: 0.9815

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.120	03/20/2017	09:30
2	CAL2	1	1	1.030	03/20/2017	09:33
3	CAL3	5	1	4.950	03/20/2017	09:36
4	CAL4	10	1	9.900	03/20/2017	09:39
5	CAL5	20	1	19.100	03/20/2017	09:42
6	CAL6	40	1	39.500	03/20/2017	09:45



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB86444

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.950	10.137	03/20/2017	09:49
2	ICB		1	0.112	0.113	03/20/2017	09:53
3	CCV1	10	1	9.910	10.096	03/20/2017	09:56
4	CCB1		1	0.116	0.118	03/20/2017	10:00
5	LB86444BL		1	0.110	0.111	03/20/2017	10:03
6	I2212-36		1	11.700	11.920	03/20/2017	10:06
7	CCV2	10	1	9.890	10.076	03/20/2017	10:09
8	CCB2		1	0.115	0.117	03/20/2017	10:12