


7470A

LB 86447
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36027



A= 0.0000e+000

B= 2.7717e-004

C= -2.6420e-002

Rho= 0.9997761

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	154	0.000	154				
0.2	0.200	0.175	-0.025	728	0.0 %	728				
2.5	2.500	2.417	-0.083	8817	0.0 %	8817				
5.0	5.000	5.162	0.162	18720	0.0 %	18720				
7.5	7.500	7.470	-0.030	27046	0.0 %	27046				
10.0	10.000	9.959	-0.041	36027	0.0 %	36027				

LB86447 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Mar 2017 11:39:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	154	-	0.0000	7470A	PPB	20 Mar 2017 11:50:27	S	Std
0.2 - 1	50.2	728	-	0.2000	7470A	PPB	20 Mar 2017 11:52:35	S	Std
2.5 - 1	52.5	8817	-	2.5000	7470A	PPB	20 Mar 2017 11:54:43	S	Std
5.0 - 1	55	18720	-	5.0000	7470A	PPB	20 Mar 2017 11:56:51	S	Std
7.5 - 1	57.5	27046	-	7.5000	7470A	PPB	20 Mar 2017 11:58:59	S	Std
10.0 - 1	510	36027	-	10.0000	7470A	PPB	20 Mar 2017 12:01:06	S	Std
ICV78 - 1	ICV78	14377	3.9584	-	7470A	PPB	20 Mar 2017 12:03:53	U	SMPL
ICB78 - 1	ICB78	150	0.0152	-	7470A	PPB	20 Mar 2017 12:05:59	U	SMPL
CCV29 - 1	CCV29	18485	5.0970	-	7470A	PPB	20 Mar 2017 12:08:05	U	SMPL
CCB29 - 1	CCB29	136	0.0113	-	7470A	PPB	20 Mar 2017 12:10:12	U	SMPL
CRA - 1	CRA	740	0.1787	-	7470A	PPB	20 Mar 2017 12:12:18	U	SMPL
HighStd - 1	HighStd	36378	10.0563	-	7470A	PPB	20 Mar 2017 12:14:25	U	SMPL
ChkStd - 1	ChkStd	25158	6.9465	-	7470A	PPB	20 Mar 2017 12:18:59	U	SMPL
PB97678BL - 1	PBW	58	-0.0103	-	7470A	PPB	20 Mar 2017 12:21:06	U	SMPL
PB97678BS - 1	LCSW	14219	3.9146	-	7470A	PPB	20 Mar 2017 12:23:13	U	SMPL
PB97678TB - 1	PB97646TB	79	-0.0045	-	7470A	PPB	20 Mar 2017 12:25:20	U	SMPL
I2277-02 - 1	S39-9003	99	0.0010	-	7470A	PPB	20 Mar 2017 12:27:29	U	SMPL
I2277-02DUP - 1	S39-9003DUP	143	0.0132	-	7470A	PPB	20 Mar 2017 12:29:37	U	SMPL
I2277-02MS - 1	S39-9003MS	13347	3.6729	-	7470A	PPB	20 Mar 2017 12:31:46	U	SMPL
I2277-02MSD - 1	S39-9003MSD	14142	3.8933	-	7470A	PPB	20 Mar 2017 12:33:52	U	SMPL
CCV30 - 1	CCV30	18562	5.1183	-	7470A	PPB	20 Mar 2017 12:35:59	U	SMPL
CCB30 - 1	CCB30	135	0.0110	-	7470A	PPB	20 Mar 2017 12:38:06	U	SMPL
I2277-02LX5 - 1		125	0.0082	-	7470A	PPB	20 Mar 2017 12:40:12	U	SMPL
I2277-02A - 1		14166	3.8999	-	7470A	PPB	20 Mar 2017 12:42:19	U	SMPL
CCV31 - 1	CCV31	18657	5.1447	-	7470A	PPB	20 Mar 2017 12:44:26	U	SMPL
CCB31 - 1	CCB31	108	0.0035	-	7470A	PPB	20 Mar 2017 12:46:33	U	SMPL