

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB86448

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP53508
calibration std. phosphate 0.5 ppm	WP53507
calibration std. phosphate 0.3 ppm	WP53505
calibration std. 0 ppm	WP53500
phosphate ICV std.	WP53511
phosphate CCV std.	WP53510
calibration std. phosphate 0.1 ppm	WP53504
calibration std. phosphate 0.05 ppm	WP53502
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP53513
Combined reagent	WP53514

Intercept: -0Slope: 0.6232Regression: 0.999952

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/16/2017	13:06
2	CAL2	0.05	1	50	50	0.032	03/16/2017	13:06
3	CAL3	0.10	1	50	50	0.065	03/16/2017	13:07
4	CAL4	0.30	1	50	50	0.184	03/20/2017	13:07
5	CAL5	0.50	1	50	50	0.309	03/16/2017	13:08
6	CAL6	1.00	1	50	50	0.625	03/16/2017	13:08



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86448

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.305	0.489	03/16/2017	13:09
2	ICB		1	50	50	0.000	0.000	03/16/2017	13:09
3	CCV1	0.50	1	50	50	0.310	0.497	03/16/2017	13:10
4	CCB1		1	50	50	0.000	0.000	03/16/2017	13:10
5	PB97630BL		1	50	50	0.000	0.000	03/16/2017	13:11
6	PB97630BS	0.50	1	50	50	0.306	0.491	03/16/2017	13:11
7	I2212-13		1	50	50	1.088	1.746	03/16/2017	13:12
8	I2278-01		1	50	50	1.621	2.601	03/16/2017	13:12
9	I2278-01DUP		1	50	50	1.620	2.599	03/16/2017	13:13
10	I2278-01MS	0.50	1	50	50	2.047	3.285	03/16/2017	13:13
11	I2278-01MSD	0.50	1	50	50	2.047	3.285	03/16/2017	13:14
12	I2212-13		5	50	50	0.520	0.834	03/16/2017	13:14
13	I2278-01		20	50	50	0.604	0.969	03/16/2017	13:30
14	I2278-01DUP		20	50	50	0.605	0.971	03/16/2017	13:30
15	CCV2	0.50	1	50	50	0.311	0.499	03/16/2017	13:31
16	CCB2		1	50	50	0.000	0.000	03/16/2017	13:31
17	I2278-01MS	0.50	20	50	50	0.615	0.987	03/16/2017	13:32
18	I2278-01MSD	0.50	20	50	50	0.613	0.984	03/16/2017	13:32
19	I2212-13		5	50	50	0.520	0.834	03/16/2017	13:33
20	CCV3	0.50	1	50	50	0.310	0.497	03/16/2017	13:33
21	CCB3		1	50	50	0.000	0.000	03/16/2017	13:34