



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86452

Temp.: 14.8 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/20/2017	12:30
2	CAL2	1	Water			7.00	03/20/2017	12:31
3	CAL3	1	Water			10.02	03/20/2017	12:32
4	ICV	1	Water			7.01	03/20/2017	12:33
5	CCV1	1	Water			2.02	03/20/2017	12:35
6	I2325-01	1	Water			4.26	03/20/2017	12:36
7	I2325-01DUP	1	Water			4.27	03/20/2017	12:37
8	CCV2	1	Water			12.01	03/20/2017	12:38

WORKLIST(Hardcopy Internal Chain)

1386452

WorkList Name : PH 12325		WorkList ID : 96670		Date : 3/20/2017 12:11:17 PM					
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2325-01	pH	Cool 4 deg C	STER15	M23	GILROY-OMEGA-CELL	03/15/2017	SM 4500-PH B

Date/Time 03/20/17 12:11 PM
 Received by: JP
 Relinquished by: JP

Date/Time 03/20/17 1:35 PM
 Received by: JP
 Relinquished by: JP