


7470A

LB86463
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35063



A= 0.0000e+000

B= 2.8627e-004

C= -4.2798e-002

Rho= 0.9998629

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.073	0.073	403	0.000	403				
0.2	0.200	0.188	-0.012	807	0.0 %	807				
2.5	2.500	2.456	-0.044	8730	0.0 %	8730				
5.0	5.000	4.908	-0.092	17293	0.0 %	17293				
7.5	7.500	7.581	0.081	26631	0.0 %	26631				
10.0	10.000	9.995	-0.005	35063	0.0 %	35063				

LB86463 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Mar 2017 10:03:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	403	-	0.0000	7470A	PPB	21 Mar 2017 10:15:08	S	Std
0.2 - 1	SO.2	807	-	0.2000	7470A	PPB	21 Mar 2017 10:17:16	S	Std
2.5 - 1	SO.5	8730	-	2.5000	7470A	PPB	21 Mar 2017 10:19:24	S	Std
5.0 - 1	SO.5	17293	-	5.0000	7470A	PPB	21 Mar 2017 10:21:33	S	Std
7.5 - 1	SO.5	26631	-	7.5000	7470A	PPB	21 Mar 2017 10:23:40	S	Std
10.0 - 1	SO	35063	-	10.0000	7470A	PPB	21 Mar 2017 10:28:17	S	Std
ICV80 - 1	ICV80	14180	4.0165	-	7470A	PPB	21 Mar 2017 10:31:06	U	SMPL
ICB80 - 1	ICB80	87	-0.0179	-	7470A	PPB	21 Mar 2017 10:33:12	U	SMPL
CCV36 - 1	CCV36	17771	5.0444	-	7470A	PPB	21 Mar 2017 10:35:19	U	SMPL
CCB36 - 1	CCB36	95	-0.0156	-	7470A	PPB	21 Mar 2017 10:37:25	U	SMPL
CRA - 1	CRA	778	0.1799	-	7470A	PPB	21 Mar 2017 10:39:31	U	SMPL
HighStd - 1	HighStd	34896	9.9467	-	7470A	PPB	21 Mar 2017 10:41:38	U	SMPL
ChkStd - 1	ChkStd	24857	7.0729	-	7470A	PPB	21 Mar 2017 10:43:45	U	SMPL
PB97707BL - 1	PBW	-2	-0.0434	-	7470A	PPB	21 Mar 2017 10:45:53	U	SMPL
PB97707BS - 1	LCSW	13823	3.9143	-	7470A	PPB	21 Mar 2017 10:48:01	U	SMPL
I2263-19 - 1	MW-103S	135	-0.0042	-	7470A	PPB	21 Mar 2017 10:50:08	U	SMPL
I2263-19DUP - 1	MW-103SDUP	206	0.0162	-	7470A	PPB	21 Mar 2017 10:52:16	U	SMPL
I2263-19MS - 1	MW-103SMS	14251	4.0368	-	7470A	PPB	21 Mar 2017 10:54:24	U	SMPL
I2263-19MSD - 1	MW-103SMSD	14072	3.9855	-	7470A	PPB	21 Mar 2017 10:56:33	U	SMPL
I2263-20 - 1	MW-103I	98	-0.0147	-	7470A	PPB	21 Mar 2017 10:58:40	U	SMPL
CCV37 - 1	CCV37	17743	5.0364	-	7470A	PPB	21 Mar 2017 11:00:49	U	SMPL
CCB37 - 1	CCB37	84	-0.0188	-	7470A	PPB	21 Mar 2017 11:02:56	U	SMPL
I2263-21 - 1	MW-103D	79	-0.0202	-	7470A	PPB	21 Mar 2017 11:05:02	U	SMPL
I2263-23 - 1	MW-104I	223	0.0210	-	7470A	PPB	21 Mar 2017 11:07:09	U	SMPL
I2263-24 - 1	MW-104D	296	0.0419	-	7470A	PPB	21 Mar 2017 11:09:16	U	SMPL
I2314-06 - 1	FIELD BLANK	362	0.0608	-	7470A	PPB	21 Mar 2017 11:11:23	U	SMPL
I2263-19LX5 - 1		344	0.0557	-	7470A	PPB	21 Mar 2017 11:13:31	U	SMPL
I2263-19A - 1		13960	3.9535	-	7470A	PPB	21 Mar 2017 11:15:39	U	SMPL
CCV38 - 1	CCV38	17528	4.9749	-	7470A	PPB	21 Mar 2017 11:17:47	U	SMPL
CCB38 - 1	CCB38	71	-0.0225	-	7470A	PPB	21 Mar 2017 11:19:55	U	SMPL