

7471A


LB86465

INSTRUMENT ID: CV8

Linear

 μ Abs.:

45592



A= 0.0000e+000

B= 2.1848e-004

C= -4.9606e-002

Rho= 0.9998548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.031	-0.031	83	0.000	83				
0.2	0.200	0.186	-0.014	1079	0.0 %	1079				
2.5	2.500	2.536	0.036	11835	0.0 %	11835				
5.0	5.000	4.984	-0.016	23037	0.0 %	23037				
7.5	7.500	7.614	0.114	35079	0.0 %	35079				
10.0	10.000	9.911	-0.089	45592	0.0 %	45592				

LB86465 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Mar 2017 12:15:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	83	-	0.00007471A	PPB		21 Mar 2017 12:34:16	S	Std
0.2 - 1	50.2	1079	-	0.20007471A	PPB		21 Mar 2017 12:36:33	S	Std
2.5 - 1	52.5	11835	-	2.50007471A	PPB		21 Mar 2017 12:38:51	S	Std
5.0 - 1	55	23037	-	5.00007471A	PPB		21 Mar 2017 12:41:08	S	Std
7.5 - 1	57.5	35079	-	7.50007471A	PPB		21 Mar 2017 12:43:25	S	Std
10.0 - 1	516	45592	-	10.00007471A	PPB		21 Mar 2017 12:45:41	S	Std
ICV81 - 1	ICV81	17990	3.8808	-7471A	PPB		21 Mar 2017 12:48:29	U	SMPL
ICB81 - 1	ICB81	107	-0.0262	-7471A	PPB		21 Mar 2017 12:50:44	U	SMPL
CCV39 - 1	CCV39	22731	4.9166	-7471A	PPB		21 Mar 2017 12:53:00	U	SMPL
CCB39 - 1	CCB39	113	-0.0249	-7471A	PPB		21 Mar 2017 12:55:15	U	SMPL
CRA - 1	CRA	1067	0.1835	-7471A	PPB		21 Mar 2017 12:57:31	U	SMPL
HighStd - 1	HighStd	45631	9.9198	-7471A	PPB		21 Mar 2017 12:59:47	U	SMPL
ChkStd - 1	ChkStd	31098	6.7447	-7471A	PPB		21 Mar 2017 13:02:03	U	SMPL
PB97709BL - 1	PBS	-20	-0.0540	-7471A	PPB		21 Mar 2017 13:04:20	U	SMPL
PB97709BS - 1	LCSS	15542	3.3460	-7471A	PPB		21 Mar 2017 13:06:37	U	SMPL
I2314-02 - 1	SGS47-9-11	716	0.1068	-7471A	PPB		21 Mar 2017 13:08:54	U	SMPL
I2314-04 - 1	FIELD DUPLICATE	462	0.0513	-7471A	PPB		21 Mar 2017 13:11:11	U	SMPL
I2331-01 - 1	FK-GF-01-002-A	3097	0.6270	-7471A	PPB		21 Mar 2017 13:13:29	U	SMPL
I2331-03 - 1	FK-GF-01-002-B	2120	0.4136	-7471A	PPB		21 Mar 2017 13:15:47	U	SMPL
I2332-01 - 1	FK-SF-01-002-A	2391	0.4728	-7471A	PPB		21 Mar 2017 13:18:03	U	SMPL
CCV40 - 1	CCV40	22694	4.9086	-7471A	PPB		21 Mar 2017 13:20:19	U	SMPL
CCB40 - 1	CCB40	122	-0.0230	-7471A	PPB		21 Mar 2017 13:22:34	U	SMPL
I2332-03 - 1	FK-SF-01-002-B	2081	0.4050	-7471A	PPB		21 Mar 2017 13:24:50	U	SMPL
I2333-01 - 1	FK-SF-01-003-A	2162	0.4227	-7471A	PPB		21 Mar 2017 13:27:06	U	SMPL
I2333-03 - 1	FK-SF-01-003-B	2900	0.5840	-7471A	PPB		21 Mar 2017 13:29:22	U	SMPL
I2333-03DUP - 1	FK-SF-01-003-BDUP	3309	0.6733	-7471A	PPB		21 Mar 2017 13:31:39	U	SMPL
I2333-03MS - 1	FK-SF-01-003-BMS	20359	4.3984	-7471A	PPB		21 Mar 2017 13:33:56	U	SMPL
I2333-03MSD - 1	FK-SF-01-003-BMSD	18690	4.0338	-7471A	PPB		21 Mar 2017 13:36:13	U	SMPL
I2333-03LX5 - 1		563	0.0734	-7471A	PPB		21 Mar 2017 13:38:30	U	SMPL
I2333-03A - 1		21370	4.6193	-7471A	PPB		21 Mar 2017 13:40:48	U	SMPL
CCV41 - 1	CCV41	22948	4.9641	-7471A	PPB		21 Mar 2017 13:43:06	U	SMPL
CCB41 - 1	CCB41	129	-0.0214	-7471A	PPB		21 Mar 2017 13:45:22	U	SMPL