



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86467

Temp.: 19.4 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/21/2017	13:15
2	CAL2	1	Water			7.02	03/21/2017	13:16
3	CAL3	1	Water			10.03	03/21/2017	13:17
4	ICV	1	Water			6.99	03/21/2017	13:18
5	CCV1	1	Water			2.02	03/21/2017	13:20
6	I2343-01	1	Water			4.39	03/21/2017	13:21
7	I2343-01DUP	1	Water			4.40	03/21/2017	13:22
8	CCV2	1	Water			12.01	03/21/2017	13:23

LB-86467

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2343

WorkList ID : 96695

Date : 3/21/2017 1:03:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2343-01	pH	Cool 4 deg C	STER07	M12	CELL-D	03/20/2017	SM 4500-PH B

Date/Time 03/21/17 11:03 PM
Received by: 71
Relinquished by: 8

Date/Time 03/21/17 21:00 PM
Received by: 08
Relinquished by: JP