

LB86471

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AV

3/21/2017 11:23

Test: Phenolics-

Sample Id	^{mg/L} Result	Dil. 1 +	Response	Errors
ICV1	0.922	0.0	0.107	
ICB1	-0.000	0.0	0.007	
CCV1	0.987	0.0	0.114	
CCB1	0.001	0.0	0.007	
PB97663BL	0.002	0.0	0.007	
PB97663BS	0.924	0.0	0.107	
I2296-05	0.008	0.0	0.008	
PB97665BL	-0.004	0.0	0.007	
PB97665BS	0.915	0.0	0.106	
I2296-01	0.010	0.0	0.008	
I2212-25	0.689	0.0	0.082	
I2296-06	0.095	0.0	0.017	
I2296-02	0.097	0.0	0.018	
CCV2	0.998	0.0	0.115	
CCB2	-0.005	0.0	0.007	

N 15
Mean 0.376
SD 0.4540
CV% 120.78

LB86471

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 21 09:47:48 2017

Tue Mar 21 11:16:48 2017

A/C

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0519	mg/l	3/21/2017 9:47:48	
0.1PPM	A	Phenolics-	P	0.1005	mg/l	3/21/2017 9:47:49	
0.25PPM	A	Phenolics-	P	0.2517	mg/l	3/21/2017 9:47:50	
0.50PPM	A	Phenolics-	P	0.4999	mg/l	3/21/2017 9:47:51	
1.0PPM	A	Phenolics-	P	0.9925	mg/l	3/21/2017 9:47:52	
2.0PPM	A	Phenolics-	P	2.0035	mg/l	3/21/2017 9:47:53	
ICV1	S	Phenolics-	P	0.9223	mg/l	3/21/2017 10:32:30	
ICB1	S	Phenolics-	P	-0.0003	mg/l	3/21/2017 10:32:31	
CCV1	S	Phenolics-	P	0.9871	mg/l	3/21/2017 10:32:32	
CCB1	S	Phenolics-	P	0.0012	mg/l	3/21/2017 10:32:33	
PB97663BL	S	Phenolics-	P	0.0017	mg/l	3/21/2017 10:32:34	
PB97663BS	S	Phenolics-	P	0.9239	mg/l	3/21/2017 10:32:35	
I2296-05	S	Phenolics-	P	0.0081	mg/l	3/21/2017 10:32:38	
PB97665BL	S	Phenolics-	P	-0.0041	mg/l	3/21/2017 10:32:40	
PB97665BS	S	Phenolics-	P	0.9147	mg/l	3/21/2017 10:32:41	
I2296-01	S	Phenolics-	P	0.0101	mg/l	3/21/2017 10:38:20	
I2212-25	S	Phenolics-	P	0.6894	mg/l	3/21/2017 11:16:43	
I2296-06	S	Phenolics-	P	0.0946	mg/l	3/21/2017 11:16:45	
I2296-02	S	Phenolics-	P	0.0969	mg/l	3/21/2017 11:16:46	
CCV2	S	Phenolics-	P	0.9984	mg/l	3/21/2017 11:16:47	
CCB2	S	Phenolics-	P	-0.0051	mg/l	3/21/2017 11:16:48	

LB86471

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

3/21/2017 10:02

Test Phenolics-


Accepted 3/21/2017 10:02

Factor 9.224

Bias 0.007

Coeff. of det. 0.999973

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.013	0.0519	0.0500	
2	Phenol-2	0.018	0.1005	0.1000	
3	Phenol-2	0.034	0.2517	0.2500	
4	Phenol-2	0.061	0.4999	0.5000	
5	Phenol-2	0.115	0.9925	1.0000	
6	Phenol-2	0.224	2.0035	2.0000	