



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86474

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	03/21/2017	17:20
2	CAL2	1	Water			7.00	03/21/2017	17:21
3	CAL3	1	Water			10.04	03/21/2017	17:22
4	ICV	1	Water			7.01	03/21/2017	17:23
5	CCV1	1	Water			2.01	03/21/2017	17:24
6	I2348-09	1	Solid	20.03	20	8.45	03/21/2017	17:25
7	I2348-09DUP	1	Solid	20.05	20	8.46	03/21/2017	17:26
8	I2349-05	1	Solid	20.03	20	6.76	03/21/2017	17:27
9	I2350-05	1	Solid	20.01	20	5.73	03/21/2017	17:30
10	I2350-10	1	Solid	20.02	20	5.86	03/21/2017	17:31
11	I2350-15	1	Solid	20.05	20	6.13	03/21/2017	17:32
12	I2350-20	1	Solid	20.04	20	6.30	03/21/2017	17:33
13	CCV2	1	Water			12.01	03/21/2017	17:35

138 5474

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2348

WorkList ID : 96703

Date : 3/21/2017 4:58:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/24/2017	Solid	I2348-09	Corrosivity	Cool 4 deg C	PSEG01	M41	JFC-SOIL-COMP	03/21/2017	9045C
03/24/2017	Solid	I2349-05	Corrosivity	Cool 4 deg C	PSEG01	M41	VNJ-212-207-COMP	03/21/2017	9045C
03/24/2017	Solid	I2350-05	Corrosivity	Cool 4 deg C	PSEG01	M41	VNJ-202-217-227-COMP	03/21/2017	9045C
03/24/2017	Solid	I2350-10	Corrosivity	Cool 4 deg C	PSEG01	M41	VNJ-238-210-221-COMP	03/21/2017	9045C
03/24/2017	Solid	I2350-15	Corrosivity	Cool 4 deg C	PSEG01	M41	72-11937-VNJ-205-COMP	03/21/2017	9045C
03/24/2017	Solid	I2350-20	Corrosivity	Cool 4 deg C	PSEG01	M41	72-11931-VNJ-226-213-COMF	03/21/2017	9045C

Date/Time 03/21/17 4:58:05 PM
 Received by: JH
 Relinquished by: _____

Date/Time 03/21/17 6:56:03
 Received by: CD
 Relinquished by: JH