



Analytical Summary Report

Analysis Method: SM4500-O G
Parameter: Dissolved Oxygen
Run Number: LB86478
BOD Water: WP53572
Starch: W2121
Conc Sulfuric Acid: W2151

Reviewed By: Indiana
Supervisor Review By: apatel
MANGANOUS SULFATE SOLUTION: W2191
Alkaline Iodide Azide: W2035
Sodium Thiosulfate, 0.025N: W2169
pH Strips: NA
Zero DO Std: WP53543

Seq	LabID	ClientID	Sample Vol (ml)	Initial Reading (mL)	Final Reading (mL)	Diff (mg/L)	Result (mg/L)
1	WINKLER1	WINKLER1	300	0.00	8.00	8.00	8.10
2	WINKLER2	WINKLER2	300	8.00	16.20	8.20	

Calibrate DO meter with average reading : 8.10

Zero DO Reading : 0.23 Calibrated to : 0.00

Seq	LabID	ClientID	DL	Sample Vol (ml)	Temperature (°C)	Result (mg/L)	Anal Date	Anal Time
1	LB86478BL	LB86478BL	1	300	19.0	0.00	03/20/2017	15:40
2	I2297-04	MW003	1	300	18.1	8.93	03/20/2017	15:43
3	I2297-04DUP	MW003DUP	1	300	18.2	8.91	03/20/2017	15:46
4	I2297-06	FB	1	300	18.8	9.08	03/20/2017	15:49
5	I2297-08	DUP001	1	300	18.5	8.32	03/20/2017	15:52

2B86478

WORKLIST(Hardcopy Internal Chain)

WorkList Name : DO032017

WorkList ID : 96674

Date : 3/20/2017 3:21:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2017	Water	I2297-04	Dissolved Oxygen	Cool 4 deg C	PWGR01	L62	MW003	03/15/2017	SM4500-O G
03/16/2017	Water	I2297-06	Dissolved Oxygen	Cool 4 deg C	PWGR01	L62	FB	03/15/2017	SM4500-O G
03/16/2017	Water	I2297-08	Dissolved Oxygen	Cool 4 deg C	PWGR01	L62	DUP001	03/15/2017	SM4500-O G

Date/Time 03/20/17 15:30
Received by: IB
Relinquished by: CP

Date/Time 03/20/17 16:25
Received by: CP
Relinquished by: IB