



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86480

Temp.: 18.3 °C

SLOPE.: 99.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/22/2017	12:15
2	CAL2	1	Water			7.00	03/22/2017	12:16
3	CAL3	1	Water			10.02	03/22/2017	12:17
4	ICV	1	Water			7.00	03/22/2017	12:18
5	CCV1	1	Water			2.02	03/22/2017	12:20
6	I2351-02	1	Water			7.74	03/22/2017	12:21
7	I2351-02DUP	1	Water			7.75	03/22/2017	12:22
8	CCV2	1	Water			12.02	03/22/2017	12:23

21386480

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	CORROSIVITY I2351	WorkList ID :	96725	Date :	3/22/2017 11:53:40 AM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/22/2017	Water	I2351-02	Corrosivity	Cool 4 deg C	PSEG01	K42	MANHOLE-001	03/21/2017	9040C

Date/Time 03/22/17 11:53 AM
Received by: 39
Relinquished by:

Date/Time 03/22/17 11:53 AM
Received by:
Relinquished by: 21