

7471A


LB 86481

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41887



A= 0.0000e+000

B= 2.3829e-004

C= 9.0268e-003

Rho= 0.9999010

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	42	0.000	42				
0.2	0.200	0.218	0.018	876	0.0 %	876				
2.5	2.500	2.492	-0.008	10421	0.0 %	10421				
5.0	5.000	4.903	-0.097	20539	0.0 %	20539				
7.5	7.500	7.577	0.077	31761	0.0 %	31761				
10.0	10.000	9.990	-0.010	41887	0.0 %	41887				

LB86481 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 22 Mar 2017 12:46:58				
Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	42	-	0.00007471A	PPB	22 Mar 2017 13:49:19	S	Std
0.2 - 1	50.2	876	-	0.20007471A	PPB	22 Mar 2017 13:51:36	S	Std
2.5 - 1	52.5	10421	-	2.50007471A	PPB	22 Mar 2017 13:53:53	S	Std
5.0 - 1	55	20539	-	5.00007471A	PPB	22 Mar 2017 13:56:11	S	Std
7.5 - 1	57.5	31761	-	7.50007471A	PPB	22 Mar 2017 14:01:17	S	Std
10.0 - 1	510	41887	-	10.00007471A	PPB	22 Mar 2017 14:03:33	S	Std
ICV83 - 1	ICV83	17001	4.0602	-7471A	PPB	22 Mar 2017 14:06:11	U	SMPL
ICB83 - 1	ICB83	-8	0.0071	-7471A	PPB	22 Mar 2017 14:08:26	U	SMPL
CCV45 - 1	CCV45	20648	4.9292	-7471A	PPB	22 Mar 2017 14:10:42	U	SMPL
CCB45 - 1	CCB45	-14	0.0057	-7471A	PPB	22 Mar 2017 14:12:57	U	SMPL
CRA - 1	CRA	847	0.2109	-7471A	PPB	22 Mar 2017 14:15:13	U	SMPL
HighStd - 1	HighStd	42305	10.0899	-7471A	PPB	22 Mar 2017 14:17:29	U	SMPL
ChkStd - 1	ChkStd	29048	6.9309	-7471A	PPB	22 Mar 2017 14:19:45	U	SMPL
PB97732BL - 1	PBS	-41	-0.0007	-7471A	PPB	22 Mar 2017 14:22:01	U	SMPL
PB97732BS - 1	LCSS	14649	3.4997	-7471A	PPB	22 Mar 2017 14:24:18	U	SMPL
I2337-01 - 1	E2-V7-BASE	4450	1.0694	-7471A	PPB	22 Mar 2017 14:26:35	U	SMPL
I2337-02 - 1	E2-W7-SSW	344	0.0910	-7471A	PPB	22 Mar 2017 14:28:52	U	SMPL
I2337-03 - 1	E2-W6-SSW	5053	1.2131	-7471A	PPB	22 Mar 2017 14:31:10	U	SMPL
I2337-04 - 1	E2-V6-BASE	1634	0.3984	-7471A	PPB	22 Mar 2017 14:33:27	U	SMPL
I2337-05 - 1	E2-Q14-ESW	43975	10.4878	-7471A	PPB	22 Mar 2017 14:35:44	U	SMPL
CCV46 - 1	CCV46	21087	5.0339	-7471A	PPB	22 Mar 2017 14:37:59	U	SMPL
CCB46 - 1	CCB46	-11	0.0064	-7471A	PPB	22 Mar 2017 14:40:15	U	SMPL
I2337-06 - 1	E2-R14-ESW	68483	16.3279	-7471A	PPB	22 Mar 2017 14:42:31	U	SMPL
I2337-07 - 1	E2-Q14-BASE	274	0.0743	-7471A	PPB	22 Mar 2017 14:44:47	U	SMPL
I2337-07DUP - 1	E2-Q14-BASEDUP	285	0.0769	-7471A	PPB	22 Mar 2017 14:47:04	U	SMPL
I2337-08 - 1	I2337-07MS	16203	3.8700	-7471A	PPB	22 Mar 2017 14:49:21	U	SMPL
I2337-09 - 1	I2337-07MSD	16321	3.8982	-7471A	PPB	22 Mar 2017 14:51:38	U	SMPL
I2337-10 - 1	E2-Q14-BASE-D	344	0.0910	-7471A	PPB	22 Mar 2017 14:53:54	U	SMPL
I2348-04 - 1	JFC-SOIL-COMP	21388	5.1056	-7471A	PPB	22 Mar 2017 14:56:12	U	SMPL
I2349-01 - 1	VNJ-212-207-COMP	5936	1.4235	-7471A	PPB	22 Mar 2017 14:58:29	U	SMPL
I2350-01 - 1	VNJ-202-217-227-COMP	1850	0.4499	-7471A	PPB	22 Mar 2017 15:00:47	U	SMPL
I2350-06 - 1	VNJ-238-210-221-COMP	5240	1.2577	-7471A	PPB	22 Mar 2017 15:03:03	U	SMPL
CCV47 - 1	CCV47	21296	5.0837	-7471A	PPB	22 Mar 2017 15:05:20	U	SMPL
CCB47 - 1	CCB47	5	0.0102	-7471A	PPB	22 Mar 2017 15:07:36	U	SMPL
I2350-11 - 1	72-11937-VNJ-205-COMP	4912	1.1795	-7471A	PPB	22 Mar 2017 15:09:53	U	SMPL
I2350-16 - 1	72-11931-VNJ-226-213-COMP	3667	0.8828	-7471A	PPB	22 Mar 2017 15:12:09	U	SMPL
I2337-07LX5 - 1		102	0.0333	-7471A	PPB	22 Mar 2017 15:14:25	U	SMPL
I2337-07A - 1		17312	4.1343	-7471A	PPB	22 Mar 2017 15:16:42	U	SMPL
I2337-05DLX2 - 1		22005	5.2526	-7471A	PPB	22 Mar 2017 15:18:58	U	SMPL
I2337-06DLX5 - 1		13565	3.2414	-7471A	PPB	22 Mar 2017 15:21:15	U	SMPL
CCV48 - 1	CCV48	20816	4.9693	-7471A	PPB	22 Mar 2017 15:23:32	U	SMPL
CCB48 - 1	CCB48	-9	0.0069	-7471A	PPB	22 Mar 2017 15:25:49	U	SMPL