

Run Number: LB86490

BalanceID: WC SC-3

DATE & TIME IN: 03/22/2017 16:10

DATE & TIME OUT: 03/22/2017 20:05

OVEN TEMP: 104 °C

OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86490BL	LB86490BL	1.5037	1.5037	100	1.5037	1.5037	1.5037	0	0
2	LB86490BS	LB86490BS	1.8592	1.8592	100	1.9139	1.9139	1.9139	0.0547	547
3	I2341-02	KY038PS027-170320	1.4007	1.4007	800	1.4031	1.4031	1.4031	0.0024	3
4	I2351-01	MANHOLE-001	1.4091	1.4091	800	1.4122	1.4122	1.4122	0.0031	3.9
5	I2352-01	DRGFB-WQ-074-BG-A	1.3994	1.3994	400	1.4176	1.4176	1.4176	0.0182	45.5
6	I2352-02	DRGFB-WQ-075-AD-A	1.4	1.4	300	1.4145	1.4145	1.4145	0.0145	48.3
7	I2352-02DUP	DRGFB-WQ-075-AD-ADUP	1.4169	1.4169	300	1.4308	1.4308	1.4308	0.0139	46.3
8	I2354-01	EFF-170321-03	1.4318	1.4318	800	1.4362	1.4362	1.4362	0.0044	5.5
9	I2355-01	032117-PRRT-E-D-M	1.418	1.418	400	1.4311	1.4311	1.4311	0.0131	32.8
10	I2355-02	032117-PRRT-E-U-M	1.401	1.401	400	1.4082	1.4082	1.4082	0.0072	18
11	I2356-01	032017-PRRT-E-U-M	1.4127	1.4127	400	1.4422	1.4422	1.4422	0.0295	73.7
12	I2356-02	032017-PRRT-E-D-M	1.4203	1.4203	400	1.4359	1.4359	1.4359	0.0156	39
13	I2361-03	OUTFALL001-170321	1.4058	1.4058	800	1.4059	1.4059	1.4059	0.0001	0.1
14	I2363-01	COMP	1.408	1.408	150	1.4476	1.4476	1.4476	0.0396	264

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

1386490

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I2363

WorkList ID : 96737

Date : 3/22/2017 3:06:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/24/2017	Water	I2341-02 -D-E	TSS	Cool 4 deg C	SHAW22	M21	KY038PS027-170320	03/20/2017	SM2540 D
	Water	I2351-01 -A	TSS	Cool 4 deg C	PSEG01	K42	MANHOLE-001	03/21/2017	SM2540 D
	Water	I2352-01 -B	TSS	Cool 4 deg C	JAYC01	M13	DRGFB-WQ-074-BG-A	03/21/2017	SM2540 D
	Water	I2352-02 -B	TSS	Cool 4 deg C	JAYC01	M13	DRGFB-WQ-075-AD-A	03/21/2017	SM2540 D
	Water	I2354-01 -E	TSS	Cool 4 deg C	VIAS01	M32	EFF-170321-03	03/21/2017	SM2540 D
	Water	I2355-01 -C	TSS	Cool 4 deg C	HDR103	M33	032117-PRRT-E-D-M	03/21/2017	SM2540 D
	Water	I2355-02 -C	TSS	Cool 4 deg C	HDR103	M33	032117-PRRT-E-U-M	03/21/2017	SM2540 D
	Water	I2356-01 -C	TSS	Cool 4 deg C	HDR103	M23	032017-PRRT-E-U-M	03/20/2017	SM2540 D
	Water	I2356-02 -C	TSS	Cool 4 deg C	HDR103	M23	032017-PRRT-E-D-M	03/20/2017	SM2540 D
	Water	I2361-03 -E	TSS	Cool 4 deg C	SHAW22	M21	OUTFALL001-170321	03/21/2017	SM2540 D
	Water	I2363-01	TSS	Cool 4 deg C	ARAM01	M21	COMP	03/22/2017	SM2540 D

Date/Time 03/22/17 3:06 PM
 Received by: JD
 Relinquished by: CP

Date/Time 03/22/17 3:28 PM
 Received by: CP
 Relinquished by: JD