



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86502

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/22/2017	17:30
2	CAL2	1	Water			7.02	03/22/2017	17:31
3	CAL3	1	Water			10.03	03/22/2017	17:32
4	ICV	1	Water			7.01	03/22/2017	17:33
5	CCV1	1	Water			2.01	03/22/2017	17:34
6	I2364-04	1	Solid	20.3	20	12.23	03/22/2017	17:35
7	I2364-04DUP	1	Solid	20.5	20	12.24	03/22/2017	17:36
8	I2364-05	1	Solid	20.2	20	7.52	03/22/2017	17:37
9	I2364-06	1	Solid	20.1	20	12.34	03/22/2017	17:39
10	CCV2	1	Water			12.02	03/22/2017	17:40

1386502

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2364 WorkList ID : 96743 Date : 3/22/2017 4:28:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2364-04	Corrosivity	Cool 4 deg C	VIAS01	M21	WC-COMP-10	03/22/2017	9045C
	Solid	I2364-05	Corrosivity	Cool 4 deg C	VIAS01	M21	WC-COMP-11	03/22/2017	9045C
	Solid	I2364-06	Corrosivity	Cool 4 deg C	VIAS01	M21	WC-COMP-12	03/22/2017	9045C

Date/Time 03/22/17 4:28 PM
 Received by: JP
 Relinquished by:

Date/Time 03/22/17 6:50 PM
 Received by: CJ
 Relinquished by: SD