



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86504

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP53724
COD CCV std, 50ppm	WP53725

Temp In (C): <u>150</u>	Date In: <u>03/23/2017</u>	Time In: <u>09:30</u>
Temp Out (C): <u>150</u>	Date Out: <u>03/23/2017</u>	Time Out: <u>11:30</u>

Intercept: -0.5466

Slope: 1.0088

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	03/16/2017	13:50
2	CAL2	10	1	Water	9.000	03/16/2017	13:50
3	CAL3	50	1	Water	50.000	03/16/2017	13:51
4	CAL4	100	1	Water	100.000	03/16/2017	13:51
5	CAL5	150	1	Water	151.000	03/16/2017	13:52



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86504

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	51.097	03/16/2017	13:52
2	ICB				1	Water	0.000	0.542	03/23/2017	12:10
3	CCV1	50			1	Water	52.000	52.088	03/23/2017	12:10
4	CCB1				1	Water	0.000	0.542	03/23/2017	12:10
5	LB86504BL				1	Water	0.000	0.542	03/23/2017	12:11
6	LB86504BS	50			1	Water	49.000	49.114	03/23/2017	12:11
7	I2351-01				1	Water	30.000	30.280	03/23/2017	12:12
8	I2351-01DUP				1	Water	29.000	29.289	03/23/2017	12:12
9	I2351-01MS	50			1	Water	76.000	75.879	03/23/2017	12:13
10	I2351-01MSD	50			1	Water	75.000	74.888	03/23/2017	12:13
11	CCV2	50			1	Water	51.000	51.097	03/23/2017	12:14
12	CCB2				1	Water	0.000	0.542	03/23/2017	12:14

2386504

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD032217 WorkList ID : 96728 Date : 3/22/2017 1:09:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/24/2017	Water	I2351-01	COD	Conc H2SO4 4c	PSEG01	K42	MANHOLE-001	03/21/2017	SM5220 D

Date/Time 03/23/17 9:05
Received by: IB
Relinquished by: CR

Date/Time 03/23/17 11:00
Received by: CR
Relinquished by: IB