


7470A

LB 86507
INSTRUMENT ID: CV4

Linear

 μ Abs.:

35848



A= 0.0000e+000

B= 2.7936e-004

C= 1.1633e-003

Rho= 0.9999706

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	24	0.000	24				
0.2	0.200	0.209	0.009	745	0.0 %	745				
2.5	2.500	2.515	0.015	8998	0.0 %	8998				
5.0	5.000	4.937	-0.063	17669	0.0 %	17669				
7.5	7.500	7.515	0.015	26898	0.0 %	26898				
10.0	10.000	10.016	0.016	35848	0.0 %	35848				

LB86507 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Mar 2017 13:09:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	30	24	-	0.00007470A	PPB	23 Mar 2017 13:56:45	S	Std
0.2 - 1	50.2	745	-	0.20007470A	PPB	23 Mar 2017 13:58:53	S	Std
2.5 - 1	52.5	8998	-	2.50007470A	PPB	23 Mar 2017 14:01:01	S	Std
5.0 - 1	55	17669	-	5.00007470A	PPB	23 Mar 2017 14:03:09	S	Std
7.5 - 1	57.5	26898	-	7.50007470A	PPB	23 Mar 2017 14:05:17	S	Std
10.0 - 1	510	35848	-	10.00007470A	PPB	23 Mar 2017 14:09:53	S	Std
ICV84 - 1	ICV84	14156	3.9557	-7470A	PPB	23 Mar 2017 14:12:27	U	SMPL
ICB84 - 1	ICB84	-72	-0.0190	-7470A	PPB	23 Mar 2017 14:14:33	U	SMPL
CCV49 - 1	CCV49	17937	5.0120	-7470A	PPB	23 Mar 2017 14:16:39	U	SMPL
CCB49 - 1	CCB49	-74	-0.0195	-7470A	PPB	23 Mar 2017 14:18:46	U	SMPL
CRA - 1	CRA	719	0.2020	-7470A	PPB	23 Mar 2017 14:20:52	U	SMPL
HighStd - 1	HighStd	35501	9.9186	-7470A	PPB	23 Mar 2017 14:22:59	U	SMPL
ChkStd - 1	ChkStd	25220	7.0466	-7470A	PPB	23 Mar 2017 14:25:06	U	SMPL
PB97746BL - 1	PBW	-83	-0.0220	-7470A	PPB	23 Mar 2017 14:27:13	U	SMPL
PB97746BS - 1	LCSW	14432	4.0329	-7470A	PPB	23 Mar 2017 14:29:21	U	SMPL
PB97746TB - 1	PB97730TB	-46	-0.0117	-7470A	PPB	23 Mar 2017 14:31:29	U	SMPL
I2348-09 - 1	JFC-SOIL-COMP	10	0.0040	-7470A	PPB	23 Mar 2017 14:33:37	U	SMPL
I2349-05 - 1	VNJ-212-207-COMP	28	0.0090	-7470A	PPB	23 Mar 2017 14:35:45	U	SMPL
I2350-05 - 1	VNJ-202-217-227-COMP	11	0.0042	-7470A	PPB	23 Mar 2017 14:37:53	U	SMPL
I2350-10 - 1	VNJ-238-210-221-COMP	30	0.0095	-7470A	PPB	23 Mar 2017 14:40:00	U	SMPL
CCV50 - 1	CCV50	17712	4.9491	-7470A	PPB	23 Mar 2017 14:42:07	U	SMPL
CCB50 - 1	CCB50	-76	-0.0201	-7470A	PPB	23 Mar 2017 14:44:13	U	SMPL
I2350-15 - 1	72-11937-VNJ-205-COMP	18	0.0062	-7470A	PPB	23 Mar 2017 14:46:20	U	SMPL
I2350-20 - 1	72-11931-VNJ-226-213-COMP	20	0.0068	-7470A	PPB	23 Mar 2017 14:48:27	U	SMPL
I2350-20DUP - 1	72-11931-VNJ-226-213-COMPDUP	19	0.0065	-7470A	PPB	23 Mar 2017 14:50:34	U	SMPL
I2350-20MS - 1	72-11931-VNJ-226-213-COMPMS	13375	3.7376	-7470A	PPB	23 Mar 2017 14:52:41	U	SMPL
I2350-20MSD - 1	72-11931-VNJ-226-213-COMPMSD	13277	3.7102	-7470A	PPB	23 Mar 2017 14:54:49	U	SMPL
PB97751BL - 1	PBW	-67	-0.0176	-7470A	PPB	23 Mar 2017 14:56:56	U	SMPL
PB97751BS - 1	LCSW	14114	3.9440	-7470A	PPB	23 Mar 2017 14:59:05	U	SMPL
I2351-01 - 1	MANHOLE-001	-41	-0.0103	-7470A	PPB	23 Mar 2017 15:01:13	U	SMPL
I2362-01 - 1	S8-0666	1	0.0014	-7470A	PPB	23 Mar 2017 15:03:22	U	SMPL
I2362-01DUP - 1	S8-0666DUP	-5	-0.0002	-7470A	PPB	23 Mar 2017 15:05:29	U	SMPL
CCV51 - 1	CCV51	17975	5.0226	-7470A	PPB	23 Mar 2017 15:07:36	U	SMPL
CCB51 - 1	CCB51	-64	-0.0167	-7470A	PPB	23 Mar 2017 15:09:42	U	SMPL
I2362-02 - 1	I2362-01MS	13582	3.7954	-7470A	PPB	23 Mar 2017 15:11:49	U	SMPL
I2362-03 - 1	I2362-01MSD	14286	3.9921	-7470A	PPB	23 Mar 2017 15:13:56	U	SMPL
I2362-04 - 1	S8-0667	753	0.2115	-7470A	PPB	23 Mar 2017 15:16:03	U	SMPL
I2362-05 - 1	S8-0663	113	0.0327	-7470A	PPB	23 Mar 2017 15:18:10	U	SMPL
I2362-06 - 1	S8-0664	80	0.0235	-7470A	PPB	23 Mar 2017 15:20:17	U	SMPL
I2362-07 - 1	S8-0662	32	0.0101	-7470A	PPB	23 Mar 2017 15:22:25	U	SMPL
I2362-08 - 1	S8-EB-2017-2	12	0.0045	-7470A	PPB	23 Mar 2017 15:24:33	U	SMPL
CCV52 - 1	CCV52	17676	4.9391	-7470A	PPB	23 Mar 2017 15:26:41	U	SMPL
CCB52 - 1	CCB52	-46	-0.0117	-7470A	PPB	23 Mar 2017 15:28:50	U	SMPL
I2350-20LX5 - 1		18	0.0062	-7470A	PPB	23 Mar 2017 15:30:57	U	SMPL
I2350-20A - 1		14509	4.0544	-7470A	PPB	23 Mar 2017 15:33:04	U	SMPL
I2362-01LX5 - 1		-38	-0.0095	-7470A	PPB	23 Mar 2017 15:35:11	U	SMPL
I2362-01A - 1		14557	4.0678	-7470A	PPB	23 Mar 2017 15:37:18	U	SMPL
CCV53 - 1	CCV53	18030	5.0380	-7470A	PPB	23 Mar 2017 15:39:25	U	SMPL
CCB53 - 1	CCB53	-66	-0.0173	-7470A	PPB	23 Mar 2017 15:41:32	U	SMPL